



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/8/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 11/05/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/06/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117184

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4455-03	EXBC8	1	1.15	8.75	9.9	7.83	76.3	
M4455-04	EXBC9	2	1.15	8.75	9.9	7.97	77.9	
M4455-05	EXBD0	3	1.15	8.75	9.9	7.96	77.8	
M4455-06	EXBD1	4	1.15	8.75	9.9	7.87	76.8	
M4455-07	EXBD2	5	1.15	8.65	9.8	7.7	75.7	
M4455-08	EXBD3	6	1.15	8.71	9.86	7.86	77.0	
M4455-09	EXBD4	7	1.15	8.75	9.9	8.04	78.7	
M4455-10	EXBD5	8	1.15	8.75	9.9	8.00	78.3	
M4455-11	EXBD6	9	1.15	8.65	9.8	8.02	79.4	
M4455-12	EXBD7	10	1.15	8.65	9.8	7.76	76.4	
M4455-13	EXBD8	11	1.15	8.79	9.94	7.96	77.5	
M4455-14	EXBD9	12	1.15	8.71	9.86	7.76	75.9	
M4455-15	EXBE0	13	1.15	8.75	9.9	7.92	77.4	
M4455-16	EXBE1	14	1.15	8.65	9.8	7.83	77.2	
M4455-17	EXBE2	15	1.15	8.75	9.9	8.00	78.3	
M4455-18	EXBE3	16	1.15	8.65	9.8	7.87	77.7	
M4455-19	EXBE4	17	1.15	8.65	9.8	7.79	76.8	
M4455-20	EXBE5	18	1.15	8.75	9.9	7.86	76.7	
M4455-21	EXBE5D	19	1.15	8.75	9.9	7.86	76.7	
M4455-22	EXBE5S	20	1.15	8.75	9.9	7.86	76.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$