



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117189

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	11/05/2021	11:30
2	CAL2	1	Water	NA	NA	20.7	7.00	11/05/2021	11:31
3	CAL3	1	Water	NA	NA	20.6	10.02	11/05/2021	11:32
4	ICV	1	Water	NA	NA	20.7	7.02	11/05/2021	11:33
5	CCV1	1	Water	NA	NA	20.6	2.01	11/05/2021	11:35
6	M4520-01	1	Solid	20.03	20	21.5	8.70	11/05/2021	11:36
7	M4537-01	1	Solid	20.04	20	20.6	9.14	11/05/2021	11:40
8	M4537-01DUP	1	Solid	20.04	20	20.07	9.15	11/05/2021	11:41
9	CCV2	1	Water	NA	NA	20.3	12.03	11/05/2021	11:45

WORKLIST(Hardcopy Internal Chain)

117189

WorkList Name : ph m4520 WorkList ID : 154015 Department : Wet-Chemistry Date : 11-05-2021 11:21:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/04/2021	Solid	M4520-01	pH	Cool 4 deg C	PSEG03	L11	40540	11/04/2021	9045D
10/22/2021	Solid	M4537-01	pH	Cool 4 deg C	EART12	D11	S1-S2-S3	10/22/2021	9045D

Date/Time 11/05/21 11:25
 Raw Sample Received by: JB Cooey
 Raw Sample Relinquished by: JB Cooey

Date/Time 11/05/21 13:00
 Raw Sample Received by: JB Cooey
 Raw Sample Relinquished by: JB Cooey