



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117190

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/05/2021	13:00
2	CAL2	1	Water	NA	NA	20.3	7.00	11/05/2021	13:01
3	CAL3	1	Water	NA	NA	20.3	10.03	11/05/2021	13:02
4	ICV	1	Water	NA	NA	20.4	7.00	11/05/2021	13:05
5	CCV1	1	Water	NA	NA	20.4	2.01	11/05/2021	13:10
6	M4557-01	1	Water	NA	NA	20.2	4.73	11/05/2021	13:13
7	M4557-01DUP	1	Water	NA	NA	20.3	4.74	11/05/2021	13:14
8	M4557-02	1	Water	NA	NA	20.2	4.81	11/05/2021	13:15
9	M4557-03	1	Water	NA	NA	20.2	4.81	11/05/2021	13:16
10	CCV2	1	Water	NA	NA	20.6	12.02	11/05/2021	13:20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4557 WorkList ID : 154020 Department : Wet-Chemistry Date : 11-05-2021 12:49:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/03/2021	Water	M4557-01	pH	Cool 4 deg C	STER13	L12	CELLS-A	11/03/2021	9040C
11/03/2021	Water	M4557-02	pH	Cool 4 deg C	STER13	L12	CELL-B	11/03/2021	9040C
11/03/2021	Water	M4557-03	pH	Cool 4 deg C	STER13	L12	CELL-C	11/03/2021	9040C

Date/Time 11/05/21 12:50
 Raw Sample Received by: JN 10/01
 Raw Sample Relinquished by: CP SM

Date/Time 11/05/21 14:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JN 10/01