



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117209

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/07/2021	08:15
2	CAL2	1	Water	NA	NA	20.4	7.00	11/07/2021	08:16
3	CAL3	1	Water	NA	NA	20.4	10.03	11/07/2021	08:17
4	ICV	1	Water	NA	NA	20.5	6.98	11/07/2021	08:20
5	CCV1	1	Water	NA	NA	20.5	2.01	11/07/2021	08:22
6	M4567-01	1	Water	NA	NA	20.3	7.74	11/07/2021	08:35
7	M4567-01DUP	1	Water	NA	NA	20.4	7.75	11/07/2021	08:36
8	M4567-02	1	Water	NA	NA	20.2	8.34	11/07/2021	08:40
9	CCV2	1	Water	NA	NA	20.3	12.03	11/07/2021	08:42

VB117209

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M4567 WorkList ID : 154050 Department : Wet-Chemistry Date : 11-07-2021 08:07:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/04/2021	Water	M4567-01	Corrosivity	Cool 4 deg C	DVIR01	L12	BASEMENT	11/04/2021	9040C
11/04/2021	Water	M4567-02	Corrosivity	Cool 4 deg C	DVIR01	L12	PUMP-ROOM	11/04/2021	9040C

Date/Time 11/07/21 08:10
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: RC (O'G'ab)

Date/Time 11/07/21 10:00
Raw Sample Received by: RC (O'G'ab)
Raw Sample Relinquished by: JD (WC)