



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/8/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 11/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117215

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4458-01	EXAY9	1	1.15	8.59	9.74	7.69	76.1	
M4458-02	EXAZ0	2	1.13	8.59	9.72	7.99	79.9	
M4458-03	EXAZ1	3	1.15	8.80	9.95	8.42	82.6	
M4458-04	EXAZ2	4	1.12	8.70	9.82	8.2	81.4	
M4458-05	EXAZ3	5	1.16	8.81	9.97	8.62	84.7	
M4458-06	EXAZ4	6	1.14	8.78	9.92	8.45	83.3	
M4458-07	EXAZ5	7	1.18	8.53	9.71	8.37	84.3	
M4458-08	EXAZ6	8	1.15	8.62	9.77	8.25	82.4	
M4458-09	EXAZ7	9	1.19	8.67	9.86	7.4	71.6	
M4458-10	EXAZ8	10	1.13	8.80	9.93	7.9	76.9	
M4458-11	EXAZ9	11	1.16	8.46	9.62	7.58	75.9	
M4458-12	EXB00	12	1.14	8.81	9.95	7.76	75.1	
M4458-13	EXB01	13	1.19	8.77	9.96	7.72	74.5	
M4458-14	EXB02	14	1.15	8.80	9.95	7.72	74.7	
M4458-15	EXB03	15	1.17	8.49	9.66	7.53	74.9	
M4458-16	EXB04	16	1.15	8.82	9.97	7.98	77.4	
M4458-17	EXB05	17	1.19	8.51	9.7	7.37	72.6	
M4458-18	EXB06	18	1.17	8.66	9.83	7.91	77.8	
M4458-19	EXB07	19	1.18	8.53	9.71	7.8	77.6	
M4458-20	EXB08	20	1.1	8.71	9.81	7.92	78.3	
M4458-21	EXB08D	21	1.1	8.71	9.81	7.92	78.3	
M4458-22	EXB08S	22	1.1	8.71	9.81	7.92	78.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$