



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/8/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 11/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117216

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4459-01	EXB09	1	1.13	8.63	9.76	7.7	76.1	
M4459-02	EXB10	2	1.13	8.77	9.9	7.89	77.1	
M4459-03	EXB11	3	1.13	8.58	9.71	7.78	77.5	
M4459-04	EXB11D	4	1.13	8.58	9.71	7.78	77.5	
M4459-05	EXB11S	5	1.13	8.58	9.71	7.78	77.5	
M4459-06	EXB12	6	1.13	8.45	9.58	7.66	77.3	
M4459-07	EXB13	7	1.15	8.83	9.98	8.08	78.5	
M4459-08	EXB14	8	1.13	8.76	9.89	8.00	78.4	
M4459-09	EXB15	9	1.13	8.39	9.52	7.7	78.3	
M4459-10	EXB16	10	1.16	8.78	9.94	8.22	80.4	
M4459-11	EXB17	11	1.13	8.49	9.62	7.76	78.1	
M4459-12	EXB18	12	1.13	8.59	9.72	8.37	84.3	
M4459-13	EXB19	13	1.14	8.82	9.96	8.26	80.7	
M4459-14	EXB20	14	1.16	8.34	9.5	7.82	79.9	
M4459-15	EXB21	15	1.15	8.70	9.85	7.67	74.9	
M4459-16	EXB22	16	1.1	8.49	9.59	7.65	77.1	
M4459-17	EXB23	17	1.13	8.67	9.8	7.7	75.8	
M4459-18	EXB24	18	1.16	8.81	9.97	8.39	82.1	
M4459-19	EXB25	19	1.13	8.56	9.69	7.93	79.4	
M4459-20	EXB26	20	1.18	8.76	9.94	7.83	75.9	
M4459-21	EXB27	21	1.15	8.83	9.98	8.66	85.1	
M4459-22	EXB28	22	1.13	8.82	9.95	8.17	79.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$