



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/8/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:30
In Date: 11/07/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 11/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117218

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4460-01	EXB29	1	1.18	8.53	9.71	7.76	77.1	
M4460-02	EXB30	2	1.18	8.51	9.69	7.58	75.2	
M4460-03	EXB31	3	1.16	8.42	9.58	7.72	77.9	
M4460-04	EXB32	4	1.17	8.54	9.71	8.04	80.4	
M4460-05	EXB33	5	1.15	8.83	9.98	8.23	80.2	
M4460-06	EXB33D	6	1.15	8.83	9.98	8.23	80.2	
M4460-07	EXB33S	7	1.15	8.83	9.98	8.23	80.2	
M4460-08	EXB34	8	1.16	8.58	9.74	7.83	77.7	
M4460-09	EXB35	9	1.15	8.40	9.55	7.57	76.4	
M4460-10	EXB36	10	1.15	8.82	9.97	8.02	77.9	
M4460-11	EXB37	11	1.14	8.42	9.56	7.65	77.3	
M4460-12	EXB38	12	1.16	8.80	9.96	8.04	78.2	
M4460-13	EXB39	13	1.13	8.51	9.64	8.2	83.1	
M4460-14	EXB40	14	1.16	8.81	9.97	8.52	83.5	
M4460-15	EXB41	15	1.13	8.76	9.89	8.53	84.5	
M4460-16	EXB42	16	1.16	8.70	9.86	8.00	78.6	
M4460-17	EXB43	17	1.18	8.64	9.82	8.00	78.9	
M4460-18	EXB44	18	1.15	8.81	9.96	8.13	79.2	
M4460-19	EXB45	19	1.12	8.69	9.81	7.87	77.7	
M4460-20	EXB46	20	1.15	8.82	9.97	8.34	81.5	
M4460-21	EXB47	21	1.14	8.80	9.94	8.28	81.1	
M4460-22	EXB48	22	1.16	8.58	9.74	7.92	78.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$