



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/9/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 11/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117224

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4461-01	EXAK6	1	1.14	8.38	9.52	7.76	79.0	
M4461-02	EXAK7	2	1.18	8.41	9.59	7.57	76.0	
M4461-03	EXAK8	3	1.15	8.65	9.8	7.96	78.7	
M4461-04	EXAK9	4	0.11	9.85	9.96	8.09	81.0	
M4461-05	EXAL0	5	1.18	8.78	9.96	8.04	78.1	
M4461-06	EXAL1	6	1.18	8.50	9.68	7.83	78.2	
M4461-07	EXAL2	7	1.14	8.44	9.58	7.72	78.0	
M4461-08	EXAL3	8	1.1	8.81	9.91	7.73	75.3	
M4461-09	EXAL4	9	1.11	8.87	9.98	8.37	81.8	
M4461-10	EXAL5	10	1.16	8.67	9.83	7.75	76.0	
M4461-11	EXAL6	11	1.19	8.43	9.62	7.81	78.5	
M4461-12	EXAL7	12	1.18	8.63	9.81	7.8	76.7	
M4461-13	EXAL7D	13	1.18	8.63	9.81	7.8	76.7	
M4461-14	EXAL7S	14	1.18	8.63	9.81	7.8	76.7	
M4461-15	EXAL8	15	1.16	8.81	9.97	8.4	82.2	
M4461-16	EXAL9	16	1.13	8.58	9.71	7.93	79.3	
M4461-17	EXAM0	17	1.18	8.56	9.74	7.7	76.2	
M4461-18	EXAM1	18	1.11	8.78	9.89	8.11	79.7	
M4461-19	EXAM2	19	1.15	8.62	9.77	7.75	76.6	
M4461-20	EXAM3	20	1.12	8.66	9.78	7.9	78.3	
M4461-21	EXAM4	21	1.19	8.59	9.78	7.83	77.3	
M4461-22	EXAM5	22	1.14	8.80	9.94	7.87	76.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$