



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/9/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:10
In Date: 11/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:01
Out Date: 11/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117225

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4462-01	EXAH6	1	1.18	8.73	9.91	8.62	85.2	
M4462-02	EXAH7	2	1.13	8.80	9.93	8.45	83.2	
M4462-03	EXAH8	3	1.13	8.64	9.77	8.62	86.7	
M4462-04	EXAH9	4	1.15	8.80	9.95	8.37	82.0	
M4462-05	EXAJ0	5	1.13	8.61	9.74	8.1	81.0	
M4462-06	EXAJ1	6	1.11	8.77	9.88	8.13	80.0	
M4462-07	EXAJ2	7	1.19	8.58	9.77	8.38	83.8	
M4462-08	EXAJ3	8	1.16	8.80	9.96	8.23	80.3	
M4462-09	EXAJ4	9	1.14	8.81	9.95	8.43	82.7	
M4462-10	EXAJ4D	10	1.14	8.81	9.95	8.43	82.7	
M4462-11	EXAJ4S	11	1.14	8.81	9.95	8.43	82.7	
M4462-12	EXAJ5	12	1.13	8.45	9.58	7.92	80.4	
M4462-13	EXAJ6	13	1.16	8.82	9.98	8.31	81.1	
M4462-14	EXAJ7	14	1.1	8.64	9.74	8.56	86.3	
M4462-15	EXAJ8	15	1.16	8.57	9.73	8.44	84.9	
M4462-16	EXAJ9	16	1.13	8.71	9.84	7.95	78.3	
M4462-17	EXAK0	17	1.11	8.86	9.97	8.04	78.2	
M4462-18	EXAK1	18	1.11	8.84	9.95	9.00	89.3	
M4462-19	EXAK2	19	1.16	8.39	9.55	7.46	75.1	
M4462-20	EXAK3	20	1.19	8.46	9.65	7.53	74.9	
M4462-21	EXAK4	21	1.16	8.82	9.98	7.83	75.6	
M4462-22	EXAK5	22	1.1	8.45	9.55	7.39	74.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$