



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/9/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 11/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 11/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117226

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4463-01	EXAG3	1	1.15	8.81	9.96	8.27	80.8	
M4463-02	EXAG4	2	1.16	8.43	9.59	7.62	76.6	
M4463-03	EXAG5	3	1.16	8.81	9.97	8.42	82.4	
M4463-04	EXAG6	4	1.16	8.78	9.94	7.96	77.4	
M4463-05	EXAG7	5	1.16	8.71	9.87	7.99	78.4	
M4463-06	EXAG8	6	1.15	8.81	9.96	7.98	77.5	
M4463-07	EXAG9	7	1.17	8.64	9.81	7.43	72.5	
M4463-08	EXAH0	8	1.18	8.79	9.97	8.05	78.2	
M4463-09	EXAH1	9	1.18	8.60	9.78	8.21	81.7	
M4463-10	EXAH2	10	1.13	8.62	9.75	8.25	82.6	
M4463-11	EXAH3	11	1.16	8.82	9.98	8.46	82.8	
M4463-12	EXAH4	12	1.11	8.50	9.61	8.13	82.6	
M4463-13	EXAH5	13	1.17	8.75	9.92	8.21	80.5	
M4463-14	EXAN6	14	1.16	8.60	9.76	7.98	79.3	
M4463-15	EXAN7	15	1.16	8.36	9.52	8.1	83.0	
M4463-16	EXAN8	16	1.16	8.75	9.91	8.62	85.3	
M4463-17	EXAN9	17	1.18	8.63	9.81	8.33	82.9	
M4463-18	EXAP0	18	1.14	8.74	9.88	8.11	79.7	
M4463-19	EXAP0D	19	1.14	8.74	9.88	8.11	79.7	
M4463-20	EXAP0S	20	1.14	8.74	9.88	8.11	79.7	
M4463-21	EXAP1	21	1.1	8.61	9.71	7.98	79.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$