



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/9/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 11/08/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 11/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117230

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4465-01	EXAM6	1	1.17	8.51	9.68	8.01	80.4	
M4465-02	EXAM7	2	1.14	8.82	9.96	8.18	79.8	
M4465-03	EXAM8	3	1.19	8.47	9.66	7.54	75.0	
M4465-04	EXAM9	4	1.13	8.64	9.77	7.86	77.9	
M4465-05	EXAN0	5	1.16	8.81	9.97	8.45	82.7	
M4465-06	EXAN1	6	1.17	8.55	9.72	7.92	78.9	
M4465-07	EXAN2	7	1.18	8.54	9.72	8.1	81.0	
M4465-08	EXAN3	8	1.18	8.46	9.64	7.87	79.1	
M4465-09	EXAN4	9	1.19	8.72	9.91	8.7	86.1	
M4465-10	EXAN5	10	1.19	8.37	9.56	7.47	75.0	
M4465-12	EXB49	11	1.18	8.54	9.72	8.15	81.6	
M4465-13	EXB50	12	1.13	8.71	9.84	7.72	75.7	
M4465-14	EXB51	13	1.13	8.55	9.68	8.13	81.9	
M4465-15	EXB52	14	1.17	8.81	9.98	8.14	79.1	
M4465-16	EXB53	15	1.17	8.50	9.67	8.48	86.0	
M4465-17	EXB54	16	1.12	8.44	9.56	7.42	74.6	
M4465-18	EXB55	17	1.18	8.33	9.51	7.39	74.5	
M4465-19	EXB56	18	1.15	8.80	9.95	7.93	77.0	
M4465-20	EXB56D	19	1.15	8.80	9.95	7.93	77.0	
M4465-21	EXB56S	20	1.15	8.80	9.95	7.93	77.0	
M4465-22	EXB57	21	1.15	8.56	9.71	7.52	74.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$