

LB117231


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38336



A= 0.0000e+000

B= 2.6298e-004

C= -3.7527e-002

Rho= 0.9997884

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.027	-0.027	41	0.000	41					12
0.2	0.200	0.224	0.024	994	0.0 %	994					-1
2.5	2.500	2.477	-0.023	9561	0.0 %	9561					2
5.0	5.000	5.115	0.115	19592	0.0 %	19592					-2
7.5	7.500	7.367	-0.133	28157	0.0 %	28157					0
10.0	10.000	10.044	0.044	38336	0.0 %	38336					

LB117231 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Nov 2021 12:25:19

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7470A	PPB	08 Nov 2021 12:36:12	S	Std
0.2 - 1	50.2	994	-	0.2000	7470A	PPB	08 Nov 2021 12:38:28	S	Std
2.5 - 1	52.5	9561	-	2.5000	7470A	PPB	08 Nov 2021 12:40:45	S	Std
5.0 - 1	55	19592	-	5.0000	7470A	PPB	08 Nov 2021 12:43:03	S	Std
7.5 - 1	57.5	28157	-	7.5000	7470A	PPB	08 Nov 2021 12:45:19	S	Std
10.0 - 1	510	38336	-	10.0000	7470A	PPB	08 Nov 2021 12:55:43	S	Std
ICV72 - 1	ICV72	15135	3.9427	-	7470A	PPB	08 Nov 2021 13:00:56	U	SMPL
ICB72 - 1	ICB72	15	-0.0336	-	7470A	PPB	08 Nov 2021 13:03:11	U	SMPL
CCV22 - 1	CCV22	19485	5.0866	-	7470A	PPB	08 Nov 2021 13:05:26	U	SMPL
CCB22 - 1	CCB22	30	-0.0296	-	7470A	PPB	08 Nov 2021 13:07:41	U	SMPL
CRA - 1	CRA	989	0.2226	-	7470A	PPB	08 Nov 2021 13:09:55	U	SMPL
HighStd - 1	HighStd	38442	10.0719	-	7470A	PPB	08 Nov 2021 13:12:11	U	SMPL
ChkStd - 1	ChkStd	26418	6.9099	-	7470A	PPB	08 Nov 2021 13:14:26	U	SMPL
PB140583BL - 1	PBW	69	-0.0194	-	7470A	PPB	08 Nov 2021 13:19:14	U	SMPL
PB140583BS - 1	LCSW	14100	3.6705	-	7470A	PPB	08 Nov 2021 13:21:29	U	SMPL
M4537-02 - 1	S1-S2-S3	-40	-0.0480	-	7470A	PPB	08 Nov 2021 13:23:45	U	SMPL
M4541-02 - 1	WC-3-AS	17	-0.0331	-	7470A	PPB	08 Nov 2021 13:26:02	U	SMPL
M4541-02DUP - 1	WC-3-ASDUP	26	-0.0307	-	7470A	PPB	08 Nov 2021 13:28:20	U	SMPL
M4541-02MS - 1	WC-3-ASMS	14384	3.7452	-	7470A	PPB	08 Nov 2021 13:30:36	U	SMPL
M4541-02MSD - 1	WC-3-ASMSD	13705	3.5666	-	7470A	PPB	08 Nov 2021 13:32:52	U	SMPL
CCV23 - 1	CCV23	19366	5.0553	-	7470A	PPB	08 Nov 2021 13:35:08	U	SMPL
CCB23 - 1	CCB23	11	-0.0346	-	7470A	PPB	08 Nov 2021 13:37:23	U	SMPL
PB140584BL - 1	PBW	24	-0.0312	-	7470A	PPB	08 Nov 2021 13:39:38	U	SMPL
PB140584BS - 1	LCSW	13578	3.5332	-	7470A	PPB	08 Nov 2021 13:41:54	U	SMPL
M4480-02 - 1	1102	31	-0.0294	-	7470A	PPB	08 Nov 2021 13:44:10	U	SMPL
M4480-03 - 1	1106	7	-0.0357	-	7470A	PPB	08 Nov 2021 13:46:25	U	SMPL
M4480-03DUP - 1	1106DUP	12	-0.0344	-	7470A	PPB	08 Nov 2021 13:48:42	U	SMPL
M4480-03MS - 1	1106MS	13376	3.4801	-	7470A	PPB	08 Nov 2021 13:50:58	U	SMPL
M4480-03MSD - 1	1106MSD	12938	3.3649	-	7470A	PPB	08 Nov 2021 13:53:14	U	SMPL
PB140586BL - 1	PBW	48	-0.0249	-	7470A	PPB	08 Nov 2021 13:58:50	U	SMPL
PB140586BS - 1	LCSW	12998	3.3807	-	7470A	PPB	08 Nov 2021 14:01:07	U	SMPL
M4523-07 - 1	FIELD-BLANK	-59	-0.0530	-	7470A	PPB	08 Nov 2021 14:03:22	U	SMPL
CCV24 - 1	CCV24	17291	4.5097	-	7470A	PPB	08 Nov 2021 14:05:37	U	SMPL
CCB24 - 1	CCB24	9	-0.0352	-	7470A	PPB	08 Nov 2021 14:07:53	U	SMPL
M4530-02 - 1	SW-P287	33	-0.0288	-	7470A	PPB	08 Nov 2021 14:10:08	U	SMPL
M4567-01 - 1	BASEMENT	76	-0.0175	-	7470A	PPB	08 Nov 2021 14:12:24	U	SMPL
M4567-02 - 1	PUMP-ROOM	1035	0.2347	-	7470A	PPB	08 Nov 2021 14:14:39	U	SMPL
M4568-01 - 1	WWTP-001	27767	7.2646	-	7470A	PPB	08 Nov 2021 14:16:55	U	SMPL
M4576-06 - 1	TW-1	-1	-0.0378	-	7470A	PPB	08 Nov 2021 14:19:11	U	SMPL
M4576-06DUP - 1	TW-1DUP	44	-0.0260	-	7470A	PPB	08 Nov 2021 14:21:29	U	SMPL
M4576-06MS - 1	TW-1MS	12695	3.3010	-	7470A	PPB	08 Nov 2021 14:31:29	U	SMPL
M4576-06MSD - 1	TW-1MSD	12634	3.2850	-	7470A	PPB	08 Nov 2021 14:33:46	U	SMPL
PB140583TB - 1	PB140544TB	-38	-0.0475	-	7470A	PPB	08 Nov 2021 14:36:02	U	SMPL
PB140584TB - 1	PB140554TB	25	-0.0310	-	7470A	PPB	08 Nov 2021 14:38:18	U	SMPL
CCV25 - 1	CCV25	17409	4.5407	-	7470A	PPB	08 Nov 2021 14:45:06	U	SMPL
CCB25 - 1	CCB25	-34	-0.0465	-	7470A	PPB	08 Nov 2021 14:47:21	U	SMPL
M4541-02LX5 - 1		22	-0.0317	-	7470A	PPB	08 Nov 2021 14:49:37	U	SMPL
M4541-02A - 1		13742	3.5763	-	7470A	PPB	08 Nov 2021 14:51:52	U	SMPL
M4480-03LX5 - 1		-26	-0.0444	-	7470A	PPB	08 Nov 2021 14:54:08	U	SMPL
M4480-03A - 1		15836	4.1270	-	7470A	PPB	08 Nov 2021 14:56:23	U	SMPL
M4576-06LX5 - 1		-53	-0.0515	-	7470A	PPB	08 Nov 2021 14:58:39	U	SMPL
M4576-06A - 1		14400	3.7494	-	7470A	PPB	08 Nov 2021 15:00:56	U	SMPL
CCV26 - 1	CCV26	18676	4.8739	-	7470A	PPB	08 Nov 2021 15:03:12	U	SMPL
CCB26 - 1	CCB26	-28	-0.0449	-	7470A	PPB	08 Nov 2021 15:05:29	U	SMPL