



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117238

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/08/2021	13:35
2	CAL2	1	Water	NA	NA	20.4	7.00	11/08/2021	13:36
3	CAL3	1	Water	NA	NA	20.4	10.03	11/08/2021	13:37
4	ICV	1	Water	NA	NA	20.4	7.01	11/08/2021	13:38
5	CCV1	1	Water	NA	NA	20.4	2.01	11/08/2021	13:40
6	M4587-01	1	Solid	20.01	20	20.7	5.44	11/08/2021	13:42
7	M4587-01DUP	1	Solid	20.03	20	20.8	5.45	11/08/2021	13:43
8	CCV2	1	Water	NA	NA	20.5	12.02	11/08/2021	13:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4582 WorkList ID : 154098 Department : Wet-Chemistry Date : 11-08-2021 13:26:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/08/2021	Solid	M4587-01	Corrosivity	Cool 4 deg C	PSEG03	L21	TOTE	11/08/2021	9045D

Date/Time 11/08/21 13:30
 Raw Sample Received by: J8 (WCI)
 Raw Sample Relinquished by: J8 (WCI)

Date/Time 11/08/21 14:30
 Raw Sample Received by: J7 (WCI)
 Raw Sample Relinquished by: J8 (WCI)