



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117239

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/08/2021	13:35
2	CAL2	1	Water	NA	NA	20.4	7.00	11/08/2021	13:36
3	CAL3	1	Water	NA	NA	20.4	10.03	11/08/2021	13:37
4	ICV	1	Water	NA	NA	20.4	7.01	11/08/2021	13:38
5	CCV1	1	Water	NA	NA	20.5	2.02	11/08/2021	15:35
6	M4586-01	1	Solid	20.02	20	20.8	5.18	11/08/2021	15:40
7	M4586-01DUP	1	Solid	20.03	20	20.9	5.19	11/08/2021	15:41
8	CCV2	1	Water	NA	NA	20.5	12.02	11/08/2021	15:42

WORKLIST(Hardcopy Internal Chain)

LB 117239

WorkList Name : PH M4488 WorkList ID : 154101 Department : Wet-Chemistry Date : 11-08-2021 15:27:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/08/2021	Solid	M4586-01	pH	Cool 4 deg C	PSEG03	L21	OH-OIL	11/08/2021	9045D

Date/Time 11/08/21 15:30
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: SP (WC)

Date/Time 11/08/21 16:30
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: SP (WC)