



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/15/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 11/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/10/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117243

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4469-01	EXAY2	1	1.13	8.74	9.87	8.34	82.5	
M4469-04	EXAR7	2	1.13	8.57	9.7	8.04	80.6	
M4469-05	EXAR8	3	1.15	8.75	9.9	8.16	80.1	
M4469-06	EXAR9	4	1.15	8.65	9.8	7.97	78.8	
M4469-07	EXBK7	5	1.13	8.67	9.8	7.82	77.2	
M4469-08	EXBK8	6	1.15	8.75	9.9	7.83	76.3	
M4469-09	EXBK9	7	1.13	8.57	9.7	7.62	75.7	
M4469-10	EXBL0	8	1.13	8.77	9.9	8.01	78.4	
M4469-11	EXBL1	9	1.13	8.67	9.8	8.15	81.0	
M4469-12	EXBL2	10	1.14	8.76	9.9	8.27	81.4	
M4469-13	EXBL3	11	1.13	8.71	9.84	7.79	76.5	
M4469-14	EXBL4	12	1.13	8.67	9.8	7.86	77.6	
M4469-15	EXBL5	13	1.13	8.74	9.87	7.98	78.4	
M4469-16	EXBL6	14	1.13	8.77	9.9	8.13	79.8	
M4469-17	EXBL7	15	1.14	8.66	9.8	7.6	74.6	
M4469-18	EXBS10	16	1.13	8.67	9.8	7.74	76.2	
M4469-19	EXBS11	17	1.13	8.83	9.96	7.8	75.5	
M4469-20	EXBS8	18	1.13	8.83	9.96	7.95	77.2	
M4469-21	EXBS8D	19	1.13	8.83	9.96	7.95	77.2	
M4469-22	EXBS8S	20	1.13	8.83	9.96	7.95	77.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$