



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/12/2021

OVENTEMP IN Celsius(°C): 108
Time IN: 14:10
In Date: 11/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:03
Out Date: 11/10/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117244

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4473-01	EXAS0	1	1.13	8.77	9.9	7.81	76.2	
M4473-02	EXAS1	2	1.15	8.65	9.8	7.85	77.5	
M4473-03	EXAS2	3	1.13	8.81	9.94	7.92	77.1	
M4473-04	EXAS3	4	1.13	8.67	9.8	7.79	76.8	
M4473-05	EXAS4	5	1.15	8.69	9.84	7.99	78.7	
M4473-06	EXAS5	6	1.15	8.75	9.9	8.62	85.4	
M4473-07	EXAS6	7	1.14	8.66	9.8	8.00	79.2	
M4473-08	EXAS7	8	1.14	8.66	9.8	8.12	80.6	
M4473-09	EXAS8	9	1.13	8.73	9.86	7.93	77.9	
M4473-10	EXAS9	10	1.14	8.76	9.9	7.68	74.7	
M4473-11	EXAT0	11	1.13	8.67	9.8	8.23	81.9	
M4473-12	EXAT1	12	1.14	8.70	9.84	7.94	78.2	
M4473-13	EXAT2	13	1.14	8.76	9.9	7.77	75.7	
M4473-14	EXAT3	14	1.13	8.83	9.96	7.84	76.0	
M4473-15	EXAT4	15	1.14	8.76	9.9	8.09	79.3	
M4473-16	EXAT5	16	1.13	8.67	9.8	8.14	80.9	
M4473-17	EXAT6	17	1.13	8.83	9.96	8.12	79.2	
M4473-18	EXAT7	18	1.14	8.76	9.9	7.44	71.9	
M4473-19	EXAT8	19	1.14	8.66	9.8	7.88	77.8	
M4473-20	EXAT8D	20	1.14	8.66	9.8	7.88	77.8	
M4473-21	EXAT8S	21	1.14	8.66	9.8	7.88	77.8	
M4473-22	EXAT9	22	1.14	8.66	9.8	8.00	79.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$