



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117251

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/09/2021	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	11/09/2021	08:16
3	CAL3	1	Water	NA	NA	20.3	10.03	11/09/2021	08:20
4	ICV	1	Water	NA	NA	20.4	7.01	11/09/2021	08:22
5	CCV1	1	Water	NA	NA	20.4	2.01	11/09/2021	08:25
6	M4593-02	1	Solid	20.03	20	20.9	8.20	11/09/2021	08:30
7	M4593-02DUP	1	Solid	20.04	20	20.8	8.21	11/09/2021	08:31
8	M4595-03	1	Solid	20.02	20	20.7	10.04	11/09/2021	08:35
9	M4595-04	1	Solid	20.01	20	20.7	10.04	11/09/2021	08:36
10	M4595-07	1	Solid	20.03	20	20.4	11.26	11/09/2021	08:40
11	M4595-08	1	Solid	20.05	20	20.5	11.26	11/09/2021	08:41
12	CCV2	1	Water	NA	NA	20.8	12.03	11/09/2021	08:43

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M4593 WorkList ID : 154118 Department : Wet-Chemistry Date : 11-09-2021 08:09:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/03/2021	Solid	M4593-02	Corrosivity	Cool 4 deg C	RMJE02	L13	WASTE-108	11/03/2021	9045D
11/05/2021	Solid	M4595-03	Corrosivity	Cool 4 deg C	ENTA05	L11	CMC-OF4-001-002	11/05/2021	9045D
11/05/2021	Solid	M4595-04	Corrosivity	Cool 4 deg C	ENTA05	L11	CMC-OF4-003-004	11/05/2021	9045D
11/05/2021	Solid	M4595-07	Corrosivity	Cool 4 deg C	ENTA05	L11	CMC-ICS-001-002	11/05/2021	9045D
11/05/2021	Solid	M4595-08	Corrosivity	Cool 4 deg C	ENTA05	L11	CMC-ICS-003-004	11/05/2021	9045D

Date/Time 11/09/21 08:10
Raw Sample Received by: JP
Raw Sample Relinquished by: JP

Date/Time 11/09/21 10:00
Raw Sample Received by: JP
Raw Sample Relinquished by: JP