

LB 117254


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33131



A= 0.0000e+000

B= 2.9780e-004

C= -2.1136e-002

Rho= 0.9996320

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.035	-0.035	-47	0.000	-47					2
0.05	0.050	0.041	-0.009	210	0.0 %	210	0				19
0.2	0.200	0.238	0.038	869	0.0 %	869					-3
2.5	2.500	2.429	-0.071	8228	0.0 %	8228					1
5.0	5.000	5.036	0.036	16983	0.0 %	16983					3
7.5	7.500	7.695	0.195	25910	0.0 %	25910					-2
10.0	10.000	9.845	-0.155	33131	0.0 %	33131					

LB117254 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 09 Nov 2021 10:42:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-47	-	0.0000	245.1	PPB	09 Nov 2021 11:03:31	S	Std
0.05 - 1	50.05	210	-	0.0500	245.1	PPB	09 Nov 2021 11:05:48	S	Std
0.2 - 1	50.2	869	-	0.2000	245.1	PPB	09 Nov 2021 11:12:48	S	Std
2.5 - 1	52.5	8228	-	2.5000	245.1	PPB	09 Nov 2021 11:15:05	S	Std
5.0 - 1	55	16983	-	5.0000	245.1	PPB	09 Nov 2021 11:17:21	S	Std
7.5 - 1	57.5	25910	-	7.5000	245.1	PPB	09 Nov 2021 11:19:37	S	Std
10.0 - 1	510	33131	-	10.0000	245.1	PPB	09 Nov 2021 11:27:06	S	Std
ICV73 - 1	ICV73	13695	4.0573	-245.1		PPB	09 Nov 2021 11:32:25	U	SMPL
ICB73 - 1	ICB73	15	-0.0167	-245.1		PPB	09 Nov 2021 11:34:40	U	SMPL
CCV27 - 1	CCV27	16152	4.7890	-245.1		PPB	09 Nov 2021 11:36:55	U	SMPL
CCB27 - 1	CCB27	12	-0.0176	-245.1		PPB	09 Nov 2021 11:39:10	U	SMPL
CRA - 1	CRA	191	0.0357	-245.1		PPB	09 Nov 2021 11:41:25	U	SMPL
HighStd - 1	HighStd	32430	9.6367	-245.1		PPB	09 Nov 2021 11:43:41	U	SMPL
ChkStd - 1	ChkStd	22686	6.7349	-245.1		PPB	09 Nov 2021 11:45:56	U	SMPL
PB140637BL - 1	PBW	181	0.0328	-245.1		PPB	09 Nov 2021 11:48:13	U	SMPL
PB140637BS - 1	LCSW	13505	4.0007	-245.1		PPB	09 Nov 2021 11:50:30	U	SMPL
M4476-02 - 1	COMP	262	0.0569	-245.1		PPB	09 Nov 2021 11:52:47	U	SMPL
M4476-02DUP - 1	COMPDUP	257	0.0554	-245.1		PPB	09 Nov 2021 11:55:04	U	SMPL
M4476-02MS - 1	COMPMS	9907	2.9292	-245.1		PPB	09 Nov 2021 11:57:20	U	SMPL
M4476-02MSD - 1	COMPMSD	9132	2.6984	-245.1		PPB	09 Nov 2021 11:59:37	U	SMPL
M4478-01 - 1	TWP02	9840	2.9093	-245.1		PPB	09 Nov 2021 12:01:53	U	SMPL
CCV28 - 1	CCV28	16711	4.9555	-245.1		PPB	09 Nov 2021 12:04:53	U	SMPL
CCB28 - 1	CCB28	40	-0.0092	-245.1		PPB	09 Nov 2021 12:07:07	U	SMPL
M4476-02LX5 - 1		50	-0.0062	-245.1		PPB	09 Nov 2021 12:09:22	U	SMPL
M4476-02A - 1		13185	3.9054	-245.1		PPB	09 Nov 2021 12:11:38	U	SMPL
CCV29 - 1	CCV29	16681	4.9465	-245.1		PPB	09 Nov 2021 12:13:53	U	SMPL
CCB29 - 1	CCB29	26	-0.0134	-245.1		PPB	09 Nov 2021 12:16:08	U	SMPL