



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117258

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	11/09/2021	09:45
2	M4586-01		1	8	100.00	>212.0	>212.0	11/09/2021	12:45
3	M4586-01DUP		1	9	100.00	>212.0	>212.0	11/09/2021	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp11921 WorkList ID : 154134 Department : Wet-Chemistry Date : 11-09-2021 11:54:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/09/2021	Solid	M4586-01	Flash Point	Cool 4 deg C	PSEG03	L21	OH-OIL	11/08/2021	1010A

Date/Time 11/09/21 11:55
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RP (Ext-105)

Date/Time 11/09/21 13:20
 Raw Sample Received by: RP (Ext-105)
 Raw Sample Relinquished by: KS (WC)