



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117259

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	11/09/2021	09:45
2	M4478-01		1	10	100.00	>212.0	>212.0	11/09/2021	10:10
3	M4478-01DUP		1	10	100.00	>212.0	>212.0	11/09/2021	10:40
4	M4480-02		1	15	100.00	>212.0	>212.0	11/09/2021	11:05
5	M4480-03		1	12	100.00	>212.0	>212.0	11/09/2021	11:35
6	M4567-01		1	13	100.00	>212.0	>212.0	11/09/2021	11:58
7	M4567-02		1	15	100.00	>212.0	>212.0	11/09/2021	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP 11821 WorkList ID : 154088 Department : Wet-Chemistry Date : 11-08-2021 10:16:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/10/2021	Water	M4478-01	Flash Point	Cool 4 deg C	loui01	TWP02	1102	11/02/2021	1010B
11/10/2021	Water	M4480-02	Flash Point	Cool 4 deg C	PSEG03	L11	1106	11/03/2021	1010B
11/10/2021	Water	M4480-03	Flash Point	Cool 4 deg C	PSEG03	L11	BASEMENT	11/03/2021	1010B
11/12/2021	Water	M4567-01	Flash Point	Cool 4 deg C	DVIR01	L12	PUMP-ROOM	11/04/2021	1010B
11/12/2021	Water	M4567-02	Flash Point	Cool 4 deg C	DVIR01	L12		11/04/2021	1010B

Date/Time 11/09/21 09:15
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: R(eat)

Date/Time 11/09/21 12:25
 Raw Sample Received by: R(eat) Lab
 Raw Sample Relinquished by: KS(wc)