



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117265

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.02	11/09/2021	13:30
2	CAL2	1	Water	NA	NA	20.4	7.00	11/09/2021	13:31
3	CAL3	1	Water	NA	NA	20.4	10.03	11/09/2021	13:32
4	ICV	1	Water	NA	NA	20.4	6.99	11/09/2021	13:35
5	CCV1	1	Water	NA	NA	20.5	2.01	11/09/2021	13:36
6	M4600-01	1	Water	NA	NA	20.6	9.08	11/09/2021	13:40
7	M4600-01DUP	1	Water	NA	NA	20.7	9.09	11/09/2021	13:41
8	CCV2	1	Water	NA	NA	20.4	12.02	11/09/2021	13:42

117265

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4600 WorkList ID : 154139 Department : Wet-Chemistry Date : 11-09-2021 13:17:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/09/2021	Water	M4600-01	pH	Cool 4 deg C	PSEG03	L11	FM786-N34145	11/09/2021	9040C

Date/Time 11/09/21 13:20
Raw Sample Received by: SP (WCC)
Raw Sample Relinquished by: CP SM

Date/Time 11/09/21 14:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP (WCC)