


LB117268

7471B

INSTRUMENT ID: CV

 μ Abs.:

49134



A= 0.0000e+000

B= 2.0531e-004

C= -1.7580e-001

Rho= 0.9998317

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.029	-0.029	717	0.000	717					15
0.2	0.200	0.230	0.030	1977	0.0 %	1977					-1
2.5	2.500	2.463	-0.037	12851	0.0 %	12851					0
5.0	5.000	4.995	-0.005	25188	0.0 %	25188					2
7.5	7.500	7.629	0.129	38014	0.0 %	38014					-1
10.0	10.000	9.912	-0.088	49134	0.0 %	49134					

LB117268 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Nov 2021 13:53:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	717	-	0.0000	7471B	PPB	09 Nov 2021 14:13:48	S	Std
0.2 - 1	S0.2	1977	-	0.2000	7471B	PPB	09 Nov 2021 14:16:05	S	Std
2.5 - 1	S2.5	12851	-	2.5000	7471B	PPB	09 Nov 2021 14:18:22	S	Std
5.0 - 1	S5	25188	-	5.0000	7471B	PPB	09 Nov 2021 14:20:39	S	Std
7.5 - 1	S7.5	38014	-	7.5000	7471B	PPB	09 Nov 2021 14:22:55	S	Std
10.0 - 1	S10	49134	-	10.0000	7471B	PPB	09 Nov 2021 14:27:47	S	Std
ICV74 - 1	ICV74	20830	4.1007	-	7471B	PPB	09 Nov 2021 14:30:30	U	SMPL
ICB74 - 1	ICB74	595	-0.0536	-	7471B	PPB	09 Nov 2021 14:32:45	U	SMPL
CCV30 - 1	CCV30	24572	4.8690	-	7471B	PPB	09 Nov 2021 14:35:00	U	SMPL
CCB30 - 1	CCB30	592	-0.0543	-	7471B	PPB	09 Nov 2021 14:37:15	U	SMPL
CRA - 1	CRA	1959	0.2264	-	7471B	PPB	09 Nov 2021 14:42:23	U	SMPL
HighStd - 1	HighStd	49665	10.0207	-	7471B	PPB	09 Nov 2021 14:44:38	U	SMPL
ChkStd - 1	ChkStd	35930	7.2009	-	7471B	PPB	09 Nov 2021 14:46:53	U	SMPL
PB140613BL - 1	PBS	511	-0.0709	-	7471B	PPB	09 Nov 2021 14:49:11	U	SMPL
PB140613BS - 1	LCSS	22244	4.3910	-	7471B	PPB	09 Nov 2021 14:51:28	U	SMPL
M4537-01 - 1	S1-S2-S3	23092	4.5651	-	7471B	PPB	09 Nov 2021 14:53:44	U	SMPL
M4541-01 - 1	WC-3-AS	80508	16.3530	-	7471B	PPB	09 Nov 2021 14:56:01	U	SMPL
M4556-01 - 1	OR-02-110521	12677	2.4269	-	7471B	PPB	09 Nov 2021 14:58:18	U	SMPL
M4562-01 - 1	T-1	7293	1.3215	-	7471B	PPB	09 Nov 2021 15:00:37	U	SMPL
M4569-01 - 1	CL-02-110521	18864	3.6971	-	7471B	PPB	09 Nov 2021 15:02:53	U	SMPL
CCV31 - 1	CCV31	24461	4.8462	-	7471B	PPB	09 Nov 2021 15:05:08	U	SMPL
CCB31 - 1	CCB31	574	-0.0580	-	7471B	PPB	09 Nov 2021 15:07:24	U	SMPL
M4574-01 - 1	VWE-B6-110521-0-10	3401	0.5224	-	7471B	PPB	09 Nov 2021 15:09:40	U	SMPL
M4574-02 - 1	VWE-B6-110521-10-20	3261	0.4937	-	7471B	PPB	09 Nov 2021 15:11:55	U	SMPL
M4574-03 - 1	VWE-B6-110521-20-30	3627	0.5688	-	7471B	PPB	09 Nov 2021 15:14:10	U	SMPL
M4576-01 - 1	SB-1-(12-12.5)	1489	0.1299	-	7471B	PPB	09 Nov 2021 15:16:26	U	SMPL
M4576-03 - 1	SB-2-(9-9.5)	197	-0.1354	-	7471B	PPB	09 Nov 2021 15:18:41	U	SMPL
M4576-04 - 1	DW-2	18986	3.7221	-	7471B	PPB	09 Nov 2021 15:20:58	U	SMPL
M4576-05 - 1	DW-3	9810	1.8383	-	7471B	PPB	09 Nov 2021 15:23:14	U	SMPL
M4577-01 - 1	SB-1	1156	0.0615	-	7471B	PPB	09 Nov 2021 15:25:31	U	SMPL
M4577-02 - 1	SB-2	1607	0.1541	-	7471B	PPB	09 Nov 2021 15:27:48	U	SMPL
M4577-03 - 1	SB-3	3695	0.5828	-	7471B	PPB	09 Nov 2021 15:30:04	U	SMPL
CCV32 - 1	CCV32	24194	4.7914	-	7471B	PPB	09 Nov 2021 15:32:19	U	SMPL
CCB32 - 1	CCB32	594	-0.0538	-	7471B	PPB	09 Nov 2021 15:34:35	U	SMPL
M4577-03DUP - 1	SB-3DUP	3873	0.6194	-	7471B	PPB	09 Nov 2021 15:36:50	U	SMPL
M4577-03MS - 1	SB-3MS	20382	4.0088	-	7471B	PPB	09 Nov 2021 15:39:06	U	SMPL
M4577-03MSD - 1	SB-3MSD	22019	4.3448	-	7471B	PPB	09 Nov 2021 15:41:21	U	SMPL
M4577-04 - 1	SB-4	1969	0.2284	-	7471B	PPB	09 Nov 2021 15:43:37	U	SMPL
PB140614BL - 1	PBS	712	-0.0296	-	7471B	PPB	09 Nov 2021 15:48:27	U	SMPL
PB140614BS - 1	LCSS	18240	3.5690	-	7471B	PPB	09 Nov 2021 15:50:42	U	SMPL
M4582-01 - 1	VWE-B10-110621-0-10	1753	0.1841	-	7471B	PPB	09 Nov 2021 15:57:53	U	SMPL
M4582-01DUP - 1	VWE-B10-110621-0-10DUP	1730	0.1794	-	7471B	PPB	09 Nov 2021 16:00:10	U	SMPL
M4582-01MS - 1	VWE-B10-110621-0-10MS	20064	3.9435	-	7471B	PPB	09 Nov 2021 16:02:27	U	SMPL
M4582-01MSD - 1	VWE-B10-110621-0-10MSD	20002	3.9307	-	7471B	PPB	09 Nov 2021 16:04:42	U	SMPL
CCV33 - 1	CCV33	24595	4.8737	-	7471B	PPB	09 Nov 2021 16:06:58	U	SMPL
CCB33 - 1	CCB33	638	-0.0448	-	7471B	PPB	09 Nov 2021 16:11:45	U	SMPL
M4582-02 - 1	VWE-B10-110621-10-20	1965	0.2276	-	7471B	PPB	09 Nov 2021 16:14:00	U	SMPL
M4582-03 - 1	VWE-B10-110621-20-30	448	-0.0838	-	7471B	PPB	09 Nov 2021 16:16:16	U	SMPL
M4583-01 - 1	VWE-B7-110621-0-10	1553	0.1430	-	7471B	PPB	09 Nov 2021 16:18:31	U	SMPL
M4583-02 - 1	VWE-B7-110621-10-20	3475	0.5376	-	7471B	PPB	09 Nov 2021 16:20:47	U	SMPL
M4583-03 - 1	VWE-B7-110621-20-30	15677	3.0428	-	7471B	PPB	09 Nov 2021 16:23:03	U	SMPL
M4595-03 - 1	CMC-OF4-001-002	917	0.0125	-	7471B	PPB	09 Nov 2021 16:25:19	U	SMPL
M4595-04 - 1	CMC-OF4-003-004	1057	0.0412	-	7471B	PPB	09 Nov 2021 16:27:35	U	SMPL
M4595-07 - 1	CMC-ICS-001-002	11274	2.1388	-	7471B	PPB	09 Nov 2021 16:29:52	U	SMPL
M4595-08 - 1	CMC-ICS-003-004	9604	1.7960	-	7471B	PPB	09 Nov 2021 16:32:09	U	SMPL
CCV34 - 1	CCV34	25047	4.9665	-	7471B	PPB	09 Nov 2021 16:34:26	U	SMPL
CCB34 - 1	CCB34	682	-0.0358	-	7471B	PPB	09 Nov 2021 16:39:25	U	SMPL
M4577-03LX5 - 1		773	-0.0171	-	7471B	PPB	09 Nov 2021 16:41:41	U	SMPL
M4577-03A - 1		22449	4.4331	-	7471B	PPB	09 Nov 2021 16:43:56	U	SMPL
M4582-01LX5 - 1		709	-0.0302	-	7471B	PPB	09 Nov 2021 16:46:12	U	SMPL
M4582-01A - 1		21789	4.2976	-	7471B	PPB	09 Nov 2021 16:48:28	U	SMPL
M4541-01DLX5 - 1		15985	3.1060	-	7471B	PPB	09 Nov 2021 16:50:43	U	SMPL
CCV35 - 1	CCV35	24466	4.8472	-	7471B	PPB	09 Nov 2021 16:52:59	U	SMPL
CCB35 - 1	CCB35	600	-0.0526	-	7471B	PPB	09 Nov 2021 16:55:15	U	SMPL