



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/10/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 11/09/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 11/10/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117274

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4601-01	AS-RO-COMP	1	1.17	8.40	9.57	8.75	90.2	
M4601-03	AS-RO-COMP-EPH-2	2	1.19	8.55	9.74	8.93	90.5	
M4602-01	R-LA-6/2-WC	3	1.15	8.82	9.97	8.00	77.7	
M4602-02	R-LA-6/2	4	1.16	8.76	9.92	5.06	44.5	
M4602-03	R-LA-6/1-WC	5	1.19	8.62	9.81	6.9	66.2	
M4602-04	R-LA-6/1	6	1.19	8.65	9.84	6.47	61.0	
M4602-05	R-LA-5/6-WC	7	1.18	8.74	9.92	6.15	56.9	
M4602-06	R-LA-5/6	8	1.11	8.44	9.55	7.16	71.7	
M4603-01	ISCO-SOIL-IDW-11082021	9	1.15	8.67	9.82	9.3	94.0	
M4604-01	TR-05-110921	18	1.14	8.49	9.63	8.96	92.1	
M4604-02	TR-05-110921-EPH-2	19	1.15	8.40	9.55	8.93	92.6	
M4605-01	VWE-B19-110821-0-10	10	1.19	8.37	9.56	9.16	95.2	
M4605-02	VWE-B19-110821-10-20	11	1.18	8.49	9.67	9.34	96.1	
M4605-03	VWE-B19-110821-20-30	12	1.17	8.36	9.53	8.55	88.3	
M4605-04	VWE-B19-110821-30-40	13	1.18	8.38	9.56	8.39	86.0	
M4605-04D	VWE-B19-110821-30-40DUP	14	1.14	8.45	9.59	8.41	86.0	
M4605-05	VWE-B19-110821-0-1	15	1.16	8.65	9.81	9.39	95.1	
M4605-06	VWE-B19-110821-20-21	16	1.12	8.46	9.58	9.26	96.2	
M4605-07	VWE-B19-110821-39-40	17	1.19	8.42	9.61	7.96	80.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$