



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117276

Slope : 98.5

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/09/2021	15:00
2	CAL2	1	Water	NA	NA	20.4	7.00	11/09/2021	15:01
3	CAL3	1	Water	NA	NA	20.4	10.03	11/09/2021	15:02
4	ICV	1	Water	NA	NA	20.5	7.00	11/09/2021	15:05
5	CCV1	1	Water	NA	NA	20.5	2.01	11/09/2021	15:06
6	M4601-01	1	Solid	20.02	20	21.5	9.01	11/09/2021	15:10
7	M4601-01DUP	1	Solid	20.03	20	21.6	9.02	11/09/2021	15:11
8	M4603-02	1	Solid	20.04	20	21.4	9.10	11/09/2021	15:15
9	CCV2	1	Water	NA	NA	20.5	12.03	11/09/2021	15:20

WORKLIST(Hardcopy Internal Chain)

WB 11/2/21

WorkList Name : CORROSIVITY M2593 WorkList ID : 154150 Department : Wet-Chemistry Date : 11-09-2021 14:11:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/09/2021	Solid	M4601-01	Corrosivity	Cool 4 deg C	PSEG03	L31	AS-RO-COMP	11/09/2021	9045D
11/08/2021	Solid	M4603-02	Corrosivity	Cool 4 deg C	PARS02	L41	ISCO-SOIL-IDW-1108202	11/08/2021	9045D

Date/Time 11/09/21 14:20
 Raw Sample Received by: J. L. (S. M.)
 Raw Sample Relinquished by: A. (S. M.)

Date/Time 11/09/21 17:00
 Raw Sample Received by: J. L. (S. M.)
 Raw Sample Relinquished by: J. L. (S. M.)