



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/11/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 11/10/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/11/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117283

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4494-01	EXBN8	1	1.15	8.59	9.74	7.8	77.4	
M4494-02	EXBN9	2	1.14	8.49	9.63	7.76	78.0	
M4494-03	EXBP0	3	1.13	8.54	9.67	7.73	77.3	
M4494-04	EXBP1	4	1.16	8.81	9.97	8.2	79.9	
M4494-05	EXBP2	5	1.12	8.58	9.7	8.05	80.8	
M4494-06	EXBP3	6	1.19	8.75	9.94	8.22	80.3	
M4494-07	EXBP4	7	1.14	8.84	9.98	8.11	78.8	
M4494-08	EXBP5	8	1.14	8.53	9.67	7.85	78.7	
M4494-09	EXBP6	9	1.18	8.60	9.78	7.9	78.1	
M4494-10	EXBP7	10	1.18	8.79	9.97	8.37	81.8	
M4494-11	EXBP8	11	1.18	8.41	9.59	7.82	79.0	
M4494-12	EXBP9	12	1.14	8.44	9.58	7.79	78.8	
M4494-13	EXBQ0	13	1.19	8.56	9.75	7.8	77.2	
M4494-14	EXBQ1	14	1.19	8.78	9.97	8.26	80.5	
M4494-15	EXBQ2	15	1.16	8.50	9.66	7.88	79.1	
M4494-16	EXBQ3	16	1.1	8.64	9.74	7.93	79.1	
M4494-17	EXBQ3D	17	1.1	8.64	9.74	7.93	79.1	
M4494-18	EXBQ3S	18	1.1	8.64	9.74	7.93	79.1	
M4494-19	EXBQ4	19	1.16	8.46	9.62	7.87	79.3	
M4494-20	EXBQ5	20	1.12	8.44	9.56	7.56	76.3	
M4494-21	EXBQ6	21	1.16	8.58	9.74	7.87	78.2	
M4494-22	EXBQ7	22	1.14	8.83	9.97	8.43	82.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$