



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/11/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:10
In Date: 11/10/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:02
Out Date: 11/11/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117284

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4496-01	EXBQ8	1	1.18	8.47	9.65	7.6	75.8	
M4496-02	EXBQ9	2	1.19	8.55	9.74	7.98	79.4	
M4496-03	EXBR0	3	1.14	8.80	9.94	8.04	78.4	
M4496-04	EXBR1	4	1.16	8.82	9.98	8.45	82.7	
M4496-05	EXBR2	5	1.16	8.70	9.86	8.92	89.2	
M4496-06	EXBR3	6	1.16	8.81	9.97	8.67	85.2	
M4496-07	EXBR4	7	1.14	8.83	9.97	7.93	76.9	
M4496-08	EXBR5	8	1.13	8.44	9.57	7.74	78.3	
M4496-09	EXBR6	9	1.16	8.78	9.94	7.88	76.5	
M4496-10	EXBR7	10	1.13	8.42	9.55	7.8	79.2	
M4496-11	EXBR8	11	1.19	8.55	9.74	8.21	82.1	
M4496-12	EXBR9	12	1.19	8.36	9.55	7.85	79.7	
M4496-13	EXBS0	13	1.18	8.76	9.94	8.02	78.1	
M4496-14	EXBS1	14	1.11	8.60	9.71	7.78	77.6	
M4496-15	EXBS2	15	1.19	8.63	9.82	8.41	83.7	
M4496-16	EXBS3	16	1.19	8.54	9.73	7.78	77.2	
M4496-17	EXBS4	17	1.16	8.42	9.58	7.71	77.8	
M4496-18	EXBS5	18	1.13	8.74	9.87	7.75	75.7	
M4496-19	EXBS6	19	1.15	8.41	9.56	7.69	77.8	
M4496-20	EXBS6D	20	1.15	8.41	9.56	7.69	77.8	
M4496-21	EXBS6S	21	1.15	8.41	9.56	7.69	77.8	
M4496-22	EXBS7	22	1.19	8.79	9.98	7.97	77.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$