



# PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 11/11/2021

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:35  
In Date: 11/10/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:25  
Out Date: 11/11/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB117287

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-----------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------|
| M4515-01 | EXBY4           | 1      | 1.17            | 8.47          | 9.64                     | 7.74                       | 77.6    |          |
| M4515-02 | EXBY5           | 2      | 1.14            | 8.63          | 9.77                     | 7.7                        | 76.0    |          |
| M4515-03 | EXBY6           | 3      | 1.15            | 8.55          | 9.7                      | 7.73                       | 77.0    |          |
| M4515-04 | EXBY7           | 4      | 1.13            | 8.61          | 9.74                     | 7.82                       | 77.7    |          |
| M4515-05 | EXBY8           | 5      | 1.15            | 8.60          | 9.75                     | 7.81                       | 77.4    |          |
| M4515-06 | EXBY9           | 6      | 1.11            | 8.47          | 9.58                     | 7.62                       | 76.9    |          |
| M4515-07 | EXBZ0           | 7      | 1.13            | 8.48          | 9.61                     | 7.73                       | 77.8    |          |
| M4515-08 | EXBZ1           | 8      | 1.18            | 8.48          | 9.66                     | 7.89                       | 79.1    |          |
| M4515-09 | EXBZ2           | 9      | 1.18            | 8.77          | 9.95                     | 8.09                       | 78.8    |          |
| M4515-10 | EXBZ3           | 10     | 1.13            | 8.45          | 9.58                     | 7.78                       | 78.7    |          |
| M4515-11 | EXBZ4           | 11     | 1.14            | 8.82          | 9.96                     | 8.18                       | 79.8    |          |
| M4515-12 | EXBZ5           | 12     | 1.13            | 8.45          | 9.58                     | 7.68                       | 77.5    |          |
| M4515-13 | EXBZ6           | 13     | 1.18            | 8.48          | 9.66                     | 7.85                       | 78.7    |          |
| M4515-14 | EXBZ7           | 14     | 1.12            | 8.76          | 9.88                     | 8.04                       | 79.0    |          |
| M4515-15 | EXBZ8           | 15     | 1.15            | 8.81          | 9.96                     | 8.07                       | 78.5    |          |
| M4515-16 | EXBZ8D          | 16     | 1.15            | 8.81          | 9.96                     | 8.07                       | 78.5    |          |
| M4515-17 | EXBZ8S          | 17     | 1.15            | 8.81          | 9.96                     | 8.07                       | 78.5    |          |
| M4515-18 | EXBZ9           | 18     | 1.11            | 8.43          | 9.54                     | 7.82                       | 79.6    |          |
| M4515-19 | EXC00           | 19     | 1.18            | 8.66          | 9.84                     | 7.8                        | 76.4    |          |
| M4515-20 | EXC03           | 20     | 1.18            | 8.57          | 9.75                     | 7.94                       | 78.9    |          |
| M4515-21 | EXC04           | 21     | 1.18            | 8.43          | 9.61                     | 7.72                       | 77.6    |          |
| M4515-22 | EXC05           | 22     | 1.15            | 8.53          | 9.68                     | 7.85                       | 78.5    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$