



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117292

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/10/2021	13:00
2	CAL2	1	Water	NA	NA	20.3	7.00	11/10/2021	13:01
3	CAL3	1	Water	NA	NA	20.3	10.03	11/10/2021	13:02
4	ICV	1	Water	NA	NA	20.4	7.02	11/10/2021	13:05
5	CCV1	1	Water	NA	NA	20.4	2.01	11/10/2021	13:09
6	M4620-01	1	Water	NA	NA	21.0	5.34	11/10/2021	13:15
7	M4620-01DUP	1	Water	NA	NA	21.1	5.35	11/10/2021	13:16
8	CCV2	1	Water	NA	NA	20.6	12.03	11/10/2021	13:20