

=====LB117296=====

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : nd Instrument ID : Konelab

11/10/2021 10:28

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.345	0.0	0.068	
ICB1	1.365	0.0	0.001	
CCV1	242.740	0.0	0.169	
CCB1	1.178	0.0	0.001	
PB140638BL	1.116	0.0	0.001	
M4482-02	9.295	0.0	0.006	
M4482-02DUP	9.302	0.0	0.006	
CCV2	247.820	0.0	0.173	
CCB2	1.396	0.0	0.001	

N	9
Mean	67.951
SD	105.1411
CV%	154.73

Aquakem v. 7.2AQ1

Results from time period:

Wed Nov 10 10:16:27 2021

Wed Nov 10 10:16:37 2021

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time
0.OPPBCN	A	Caynide-Free	P	0.9434	µg/l	11/10/2021 8:55:00
5.OPPBCN	A	Caynide-Free	P	6.0381	µg/l	11/10/2021 8:55:01
10PPBCN	A	Caynide-Free	P	10.6835	µg/l	11/10/2021 8:55:02
50PPBCN	A	Caynide-Free	P	47.6222	µg/l	11/10/2021 8:55:03
100PPBCN	A	Caynide-Free	P	99.0143	µg/l	11/10/2021 8:55:04
250PPBCN	A	Caynide-Free	P	250.575	µg/l	11/10/2021 8:55:05
500PPBCN	A	Caynide-Free	P	500.1234	µg/l	11/10/2021 8:55:06
ICV1	S	Caynide-Free	P	97.3452	µg/l	11/10/2021 10:16:27
ICB1	S	Caynide-Free	P	1.3649	µg/l	11/10/2021 10:16:28
CCV1	S	Caynide-Free	P	242.7403	µg/l	11/10/2021 10:16:29
CCB1	S	Caynide-Free	P	1.1777	µg/l	11/10/2021 10:16:30
PB140638BL	S	Caynide-Free	P	1.1165	µg/l	11/10/2021 10:16:31
M4482-02	S	Caynide-Free	P	9.2945	µg/l	11/10/2021 10:16:33
M4482-02DUP	S	Caynide-Free	P	9.302	µg/l	11/10/2021 10:16:34
CCV2	S	Caynide-Free	P	247.8196	µg/l	11/10/2021 10:16:36
CCB2	S	Caynide-Free	P	1.396	µg/l	11/10/2021 10:16:37

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

11/10/2021 8:56 Reviewed by : pl Instrument ID : Konelab


Test Total CN

Accepted 11/10/2021 8:56

Factor 1435
Bias 0

Coeff. of det. 0.999954

Errors



	Calibrator	Response	Calc. con.	Conc.	Rel. Errors
1	0.0PPBCN	0.001	0.9434	0.0000	
2	5.0PPBCN	0.004	6.0381	5.0000	20.8
3	10PPBCN	0.007	10.6835	10.0000	6.8
4	50PPBCN	0.033	47.6222	50.0000	-4.8
5	100PPBCN	0.069	99.0143	100.0000	-1.0
6	250PPBCN	0.175	250.5750	250.0000	0.2
7	500PPBCN	0.348	500.1234	500.0000	0.0

AB
11/10/21