

Analysis Method: SM2510 BANALYST: apatelParameter: ConductanceSUPERVISOR REVIEW BY: ketankumarRun Number: LB117302Meter ID: WC Conductance Meter-1Thermometer ID: WC pH meter-1

Reagent/Standard	Lot/Log #
conductivity standard	WP92563

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB117302BL		1	21.2	0.72	0.78	11/10/2021	14:05
2	LB117302BS	1413	1	21.5	1315.00	1409.21	11/10/2021	14:08
3	M4620-01		1	21.6	5.09	5.44	11/10/2021	14:12
4	M4620-01DUP		1	21.6	5.08	5.43	11/10/2021	14:15

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Conductance-111021 WorkList ID : 154198 Department : Wet-Chemistry Date : 11-10-2021 13:13:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/23/2021	Water	M4620-01	Conductance	Cool 4 deg C	STER06	D11	POOL-WATER	11/04/2021	SM2510 B

Date/Time 11/10/21 14:00
 Raw Sample Received by: AD (w/c)
 Raw Sample Relinquished by: JD (w/c)

Date/Time 11/10/21 14:00
 Raw Sample Received by: JD (w/c)
 Raw Sample Relinquished by: AD (w/c)