

LB117303


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50541



A= 0.0000e+000

B= 1.9705e-004

C= -6.3250e-002

Rho= 0.9998067

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/0/1
0.0	0.000	-0.056	-0.056	35	0.000	35					15
0.2	0.200	0.230	0.030	1486	0.0 %	1486					-1
2.5	2.500	2.487	-0.013	12943	0.0 %	12943					0
5.0	5.000	5.017	0.017	25781	0.0 %	25781					2
7.5	7.500	7.627	0.127	39029	0.0 %	39029					-1
10.0	10.000	9.896	-0.104	50541	0.0 %	50541					

LB117303 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Nov 2021 15:17:46

Sample ID	Extended ID	μ Abs.	Conc.	Std. Cond. Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.00007471B	PPB	10 Nov 2021 15:51:27	S	Std
0.2 - 1	50.2	1486	-	0.20007471B	PPB	10 Nov 2021 15:53:44	S	Std
2.5 - 1	52.5	12943	-	2.50007471B	PPB	10 Nov 2021 15:56:02	S	Std
5.0 - 1	55	25781	-	5.00007471B	PPB	10 Nov 2021 15:58:18	S	Std
7.5 - 1	57.5	39029	-	7.50007471B	PPB	10 Nov 2021 16:00:36	S	Std
10.0 - 1	57.5	50541	-	10.00007471B	PPB	10 Nov 2021 16:05:32	S	Std
ICV77 - 1	ICV77	21084	4.0913	-7471B	PPB	10 Nov 2021 16:08:16	U	SMPL
ICB77 - 1	ICB77	72	-0.0491	-7471B	PPB	10 Nov 2021 16:13:03	U	SMPL
CCV42 - 1	CCV42	24891	4.8414	-7471B	PPB	10 Nov 2021 16:15:18	U	SMPL
CCB42 - 1	CCB42	237	-0.0165	-7471B	PPB	10 Nov 2021 16:20:03	U	SMPL
CRA - 1	CRA	1425	0.2175	-7471B	PPB	10 Nov 2021 16:22:18	U	SMPL
HighStd - 1	HighStd	50589	9.9051	-7471B	PPB	10 Nov 2021 16:24:33	U	SMPL
ChkStd - 1	ChkStd	34688	6.7719	-7471B	PPB	10 Nov 2021 16:26:48	U	SMPL
PB140677BL - 1	PBS	248	-0.0144	-7471B	PPB	10 Nov 2021 16:31:35	U	SMPL
PB140677BS - 1	LCSS	20425	3.9614	-7471B	PPB	10 Nov 2021 16:33:51	U	SMPL
M4068-01 - 1	LOD-MDL-SOIL-01-QT4-2021	526	0.0404	-7471B	PPB	10 Nov 2021 16:38:38	U	SMPL
M4068-02 - 1	LOQ-SOIL-02-QT4-2021	1315	0.1959	-7471B	PPB	10 Nov 2021 16:40:54	U	SMPL
M4601-01 - 1	AS-RO-COMP	1804	0.2922	-7471B	PPB	10 Nov 2021 16:43:11	U	SMPL
M4602-01 - 1	R-LA-6/2-WC	3750	0.6757	-7471B	PPB	10 Nov 2021 16:45:28	U	SMPL
M4602-03 - 1	R-LA-6/1-WC	7584	1.4311	-7471B	PPB	10 Nov 2021 16:47:43	U	SMPL
CCV43 - 1	CCV43	24823	4.8280	-7471B	PPB	10 Nov 2021 16:49:58	U	SMPL
CCB43 - 1	CCB43	178	-0.0282	-7471B	PPB	10 Nov 2021 16:54:44	U	SMPL
M4602-05 - 1	R-LA-5/6-WC	9477	1.8042	-7471B	PPB	10 Nov 2021 16:56:59	U	SMPL
M4603-01 - 1	ISCO-SOIL-IDW-11082021	4448	0.8132	-7471B	PPB	10 Nov 2021 16:59:14	U	SMPL
M4604-01 - 1	TR-05-110921	1488	0.2300	-7471B	PPB	10 Nov 2021 17:01:29	U	SMPL
M4605-01 - 1	VWE-B19-110821-0-10	494	0.0341	-7471B	PPB	10 Nov 2021 17:03:45	U	SMPL
M4605-01DUP - 1	VWE-B19-110821-0-10DUP	546	0.0443	-7471B	PPB	10 Nov 2021 17:06:00	U	SMPL
M4605-01MS - 1	VWE-B19-110821-0-10MS	18990	3.6786	-7471B	PPB	10 Nov 2021 17:08:17	U	SMPL
M4605-01MSD - 1	VWE-B19-110821-0-10MSD	20135	3.9043	-7471B	PPB	10 Nov 2021 17:10:33	U	SMPL
M4605-02 - 1	VWE-B19-110821-10-20	457	0.0268	-7471B	PPB	10 Nov 2021 17:12:51	U	SMPL
M4605-03 - 1	VWE-B19-110821-20-30	321	0.0000	-7471B	PPB	10 Nov 2021 17:15:08	U	SMPL
M4605-04 - 1	VWE-B19-110821-30-40	93	-0.0449	-7471B	PPB	10 Nov 2021 17:17:23	U	SMPL
CCV44 - 1	CCV44	24935	4.8501	-7471B	PPB	10 Nov 2021 17:19:39	U	SMPL
CCB44 - 1	CCB44	265	-0.0110	-7471B	PPB	10 Nov 2021 17:21:54	U	SMPL
M4605-01LX5 - 1		115	-0.0406	-7471B	PPB	10 Nov 2021 17:24:10	U	SMPL
M4605-01A - 1		21062	4.0869	-7471B	PPB	10 Nov 2021 17:26:25	U	SMPL
CCV45 - 1	CCV45	24989	4.8607	-7471B	PPB	10 Nov 2021 17:28:41	U	SMPL
CCB45 - 1	CCB45	204	-0.0231	-7471B	PPB	10 Nov 2021 17:30:56	U	SMPL