

LB117311


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35442



A= 0.0000e+000

B= 2.8559e-004

C= -6.1205e-002

Rho= 0.9995958

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.058	-0.058	10	0.000	10					-
0.2	0.200	0.236	0.036	1041	0.0 %	1041					18
2.5	2.500	2.505	0.005	8986	0.0 %	8986					0
5.0	5.000	5.147	0.147	18235	0.0 %	18235					3
7.5	7.500	7.310	-0.190	25810	0.0 %	25810					-3
10.0	10.000	10.061	0.061	35442	0.0 %	35442					1

LB117311 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Nov 2021 09:48:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	7470A	PPB	11 Nov 2021 10:32:09	S	Std
0.2 - 1	50.2	1041	-	0.2000	7470A	PPB	11 Nov 2021 10:37:12	S	Std
2.5 - 1	52.5	8986	-	2.5000	7470A	PPB	11 Nov 2021 10:39:30	S	Std
5.0 - 1	55	18235	-	5.0000	7470A	PPB	11 Nov 2021 10:41:47	S	Std
7.5 - 1	57.5	25810	-	7.5000	7470A	PPB	11 Nov 2021 10:49:08	S	Std
10.0 - 1	510	35442	-	10.0000	7470A	PPB	11 Nov 2021 10:56:34	S	Std
ICV78 - 1	ICV78	14381	4.0459	-7470A	PPB	PPB	11 Nov 2021 10:59:23	U	SMPL
ICB78 - 1	ICB78	138	-0.0218	-7470A	PPB	PPB	11 Nov 2021 11:06:37	U	SMPL
CCV46 - 1	CCV46	17439	4.9192	-7470A	PPB	PPB	11 Nov 2021 11:08:51	U	SMPL
CCB46 - 1	CCB46	81	-0.0381	-7470A	PPB	PPB	11 Nov 2021 11:13:36	U	SMPL
CRA - 1	CRA	1016	0.2290	-7470A	PPB	PPB	11 Nov 2021 11:15:51	U	SMPL
HighStd - 1	HighStd	35036	9.9448	-7470A	PPB	PPB	11 Nov 2021 11:18:06	U	SMPL
ChkStd - 1	ChkStd	24780	7.0157	-7470A	PPB	PPB	11 Nov 2021 11:20:21	U	SMPL
PB140675BL - 1	PBW	44	-0.0486	-7470A	PPB	PPB	11 Nov 2021 11:25:29	U	SMPL
PB140675BS - 1	LCSW	13382	3.7606	-7470A	PPB	PPB	11 Nov 2021 11:27:44	U	SMPL
M4593-02 - 1	WASTE-108	-43	-0.0735	-7470A	PPB	PPB	11 Nov 2021 11:30:01	U	SMPL
M4601-02 - 1	AS-RO-COMP	1	-0.0609	-7470A	PPB	PPB	11 Nov 2021 11:32:17	U	SMPL
M4601-02DUP - 1	AS-RO-COMPDUP	0	-0.0612	-7470A	PPB	PPB	11 Nov 2021 11:34:34	U	SMPL
M4601-02MS - 1	AS-RO-COMPMS	13239	3.7197	-7470A	PPB	PPB	11 Nov 2021 11:36:51	U	SMPL
M4601-02MSD - 1	AS-RO-COMPMSD	13490	3.7914	-7470A	PPB	PPB	11 Nov 2021 11:39:06	U	SMPL
CCV47 - 1	CCV47	17817	5.0272	-7470A	PPB	PPB	11 Nov 2021 11:41:21	U	SMPL
CCB47 - 1	CCB47	101	-0.0324	-7470A	PPB	PPB	11 Nov 2021 11:43:37	U	SMPL
PB140676BL - 1	PBW	132	-0.0235	-7470A	PPB	PPB	11 Nov 2021 11:45:52	U	SMPL
PB140676BS - 1	LCSW	12693	3.5638	-7470A	PPB	PPB	11 Nov 2021 11:48:07	U	SMPL
M4068-07 - 1	LOD-MDL-WATER-01-QT4-2021	745	0.1516	-7470A	PPB	PPB	11 Nov 2021 11:52:54	U	SMPL
M4068-09 - 1	MDL-WATER-03-QT4-2021	433	0.0625	-7470A	PPB	PPB	11 Nov 2021 12:00:14	U	SMPL
M4605-08 - 1	VWE-B19-110821-GW	766	0.1576	-7470A	PPB	PPB	11 Nov 2021 12:07:18	U	SMPL
M4605-08DUP - 1	VWE-B19-110821-GWDUP	701	0.1390	-7470A	PPB	PPB	11 Nov 2021 12:09:34	U	SMPL
M4605-08MS - 1	VWE-B19-110821-GWMS	16006	4.5100	-7470A	PPB	PPB	11 Nov 2021 12:11:51	U	SMPL
M4605-08MSD - 1	VWE-B19-110821-GWMSD	15654	4.4094	-7470A	PPB	PPB	11 Nov 2021 12:14:08	U	SMPL
M4606-23 - 1	FB-02-20211109	-35	-0.0712	-7470A	PPB	PPB	11 Nov 2021 12:16:25	U	SMPL
PB140675TB - 1	PB140642TB	9	-0.0586	-7470A	PPB	PPB	11 Nov 2021 12:18:41	U	SMPL
CCV48 - 1	CCV48	17913	5.0546	-7470A	PPB	PPB	11 Nov 2021 12:20:56	U	SMPL
CCB48 - 1	CCB48	152	-0.0178	-7470A	PPB	PPB	11 Nov 2021 12:32:50	U	SMPL
M4601-02LX5 - 1		22	-0.0549	-7470A	PPB	PPB	11 Nov 2021 12:35:05	U	SMPL
M4601-02A - 1		14263	4.0122	-7470A	PPB	PPB	11 Nov 2021 12:37:20	U	SMPL
M4605-08LX5 - 1		174	-0.0115	-7470A	PPB	PPB	11 Nov 2021 12:44:53	U	SMPL
M4605-08A - 1		12198	3.4224	-7470A	PPB	PPB	11 Nov 2021 12:55:16	U	SMPL
CCV49 - 1	CCV49	16362	4.6116	-7470A	PPB	PPB	11 Nov 2021 12:57:32	U	SMPL
CCB49 - 1	CCB49	62	-0.0435	-7470A	PPB	PPB	11 Nov 2021 12:59:48	U	SMPL