

LB117318


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47417



A= 0.0000e+000

B= 2.0916e-004

C= -5.9279e-003

Rho= 0.9998437

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.023	-0.023	-80	0.000	-80					-
0.2	0.200	0.218	0.018	1069	0.0 %	1069					9
2.5	2.500	2.472	-0.028	11846	0.0 %	11846					-1
5.0	5.000	4.995	-0.005	23909	0.0 %	23909					0
7.5	7.500	7.627	0.127	36493	0.0 %	36493					2
10.0	10.000	9.912	-0.088	47417	0.0 %	47417					-1

LB117318 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 11 Nov 2021 13:59:48				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-80	-	0.00007471B	PPB	11 Nov 2021 14:17:05	S	Std
0.2 - 1	50.2	1069	-	0.20007471B	PPB	11 Nov 2021 14:21:55	S	Std
2.5 - 1	52.5	11846	-	2.50007471B	PPB	11 Nov 2021 14:24:12	S	Std
5.0 - 1	55	23909	-	5.00007471B	PPB	11 Nov 2021 14:26:29	S	Std
7.5 - 1	57.5	36493	-	7.50007471B	PPB	11 Nov 2021 14:28:45	S	Std
10.0 - 1	510	47417	-	10.00007471B	PPB	11 Nov 2021 14:31:01	S	Std
ICV79 - 1	ICV79	19825	4.1406	-7471B	PPB	11 Nov 2021 14:33:46	U	SMPL
ICB79 - 1	ICB79	-76	-0.0218	-7471B	PPB	11 Nov 2021 14:36:01	U	SMPL
CCV50 - 1	CCV50	24195	5.0546	-7471B	PPB	11 Nov 2021 14:38:17	U	SMPL
CCB50 - 1	CCB50	-82	-0.0231	-7471B	PPB	11 Nov 2021 14:40:31	U	SMPL
CRA - 1	CRA	1086	0.2212	-7471B	PPB	11 Nov 2021 14:42:47	U	SMPL
HighStd - 1	HighStd	47631	9.9564	-7471B	PPB	11 Nov 2021 14:45:02	U	SMPL
ChkStd - 1	ChkStd	34898	7.2932	-7471B	PPB	11 Nov 2021 14:49:48	U	SMPL
PB140710BL - 1	PBS	-118	-0.0306	-7471B	PPB	11 Nov 2021 14:52:04	U	SMPL
PB140710BS - 1	LCSS	17478	3.6497	-7471B	PPB	11 Nov 2021 14:54:21	U	SMPL
M4068-01 - 1	LOD-MDL-SOIL-01-QT4-2021	768	0.1547	-7471B	PPB	11 Nov 2021 15:03:02	U	SMPL
M4068-03 - 1	MDL-SOIL-03-QT4-2021	282	0.0531	-7471B	PPB	11 Nov 2021 15:07:52	U	SMPL
M4386-11 - 1	PT-MET-SOIL	257479	53.8475	-7471B	PPB	11 Nov 2021 15:10:09	U	SMPL
M4606-14 - 1	OF-3-20211109	10338	2.1563	-7471B	PPB	11 Nov 2021 15:17:51	U	SMPL
M4606-14DUP - 1	OF-3-20211109DUP	10235	2.1348	-7471B	PPB	11 Nov 2021 15:20:06	U	SMPL
CCV51 - 1	CCV51	23992	5.0122	-7471B	PPB	11 Nov 2021 15:22:22	U	SMPL
CCB51 - 1	CCB51	-59	-0.0183	-7471B	PPB	11 Nov 2021 15:24:37	U	SMPL
M4606-15 - 1	M4606-14MS	30223	6.3154	-7471B	PPB	11 Nov 2021 15:26:52	U	SMPL
M4606-16 - 1	M4606-14MSD	32798	6.8540	-7471B	PPB	11 Nov 2021 15:29:07	U	SMPL
M4606-17 - 1	OF-4-20211109	7927	1.6521	-7471B	PPB	11 Nov 2021 15:31:24	U	SMPL
M4606-21 - 1	DUP-03-20211109	9060	1.8890	-7471B	PPB	11 Nov 2021 15:33:40	U	SMPL
M4612-01 - 1	GSMP-EB-ONLY	853	0.1725	-7471B	PPB	11 Nov 2021 15:35:56	U	SMPL
M4619-01 - 1	HR-01-111021	829	0.1675	-7471B	PPB	11 Nov 2021 15:38:13	U	SMPL
M4619-03 - 1	HR-02-111021	1806	0.3718	-7471B	PPB	11 Nov 2021 15:40:29	U	SMPL
M4625-01 - 1	SU-02-111021	1879	0.3871	-7471B	PPB	11 Nov 2021 15:42:46	U	SMPL
M4625-03 - 1	SU-03-111021	1509	0.3097	-7471B	PPB	11 Nov 2021 15:45:03	U	SMPL
M4628-01 - 1	LAMINGTON-COMP	2363	0.4883	-7471B	PPB	11 Nov 2021 15:47:18	U	SMPL
CCV52 - 1	CCV52	23873	4.9873	-7471B	PPB	11 Nov 2021 15:49:34	U	SMPL
CCB52 - 1	CCB52	-46	-0.0155	-7471B	PPB	11 Nov 2021 15:51:50	U	SMPL
M4635-01 - 1	VWE-B20-11921-0-10	265	0.0495	-7471B	PPB	11 Nov 2021 15:54:06	U	SMPL
M4606-14LX5 - 1		1362	0.2789	-7471B	PPB	11 Nov 2021 16:03:19	U	SMPL
M4606-14A - 1		30536	6.3809	-7471B	PPB	11 Nov 2021 16:05:41	U	SMPL
M4386-11DLX10 - 1	PT-MET-SOIL	27914	5.8325	-7471B	PPB	11 Nov 2021 16:14:49	U	SMPL
CCV53 - 1	CCV53	24336	5.0841	-7471B	PPB	11 Nov 2021 16:17:05	U	SMPL
CCB53 - 1	CCB53	-55	-0.0174	-7471B	PPB	11 Nov 2021 16:19:21	U	SMPL