



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117319

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	11/11/2021	12:58
2	M4600-01		1	10	100.00	>212.0	>212.0	11/11/2021	13:20
3	M4600-01DUP		1	10	100.00	>212.0	>212.0	11/11/2021	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117319

WorkList Name : FP111121 WorkList ID : 154223 Department : Wet-Chemistry Date : 11-11-2021 08:53:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/12/2021	Water	M4600-01	Flash Point	Cool 4 deg C	PSEG03	L11	FM786-N34145	11/09/2021	1010B

Date/Time 11/11/21 12:40
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JB (DC)

Date/Time 11/11/21 13:50
 Raw Sample Received by: JB (DC)
 Raw Sample Relinquished by: KS (WC)