



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117329

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	11/11/2021	16:30
2	CAL2	1	Water	NA	NA	20.8	7.01	11/12/2021	16:31
3	CAL3	1	Water	NA	NA	20.8	10.03	11/11/2021	16:32
4	ICV	1	Water	NA	NA	20.8	6.99	11/11/2021	16:35
5	CCV1	1	Water	NA	NA	20.7	2.01	11/11/2021	16:40
6	M4656-06	1	Water	NA	NA	20.4	6.82	11/11/2021	16:43
7	M4656-07	1	Water	NA	NA	20.5	9.12	11/11/2021	16:45
8	M4656-08	1	Water	NA	NA	20.5	6.28	11/11/2021	16:46
9	M4656-09	1	Water	NA	NA	20.6	11.87	11/11/2021	16:50
10	M4656-09DUP	1	Water	NA	NA	20.7	11.88	11/11/2021	16:51
11	CCV2	1	Water	NA	NA	20.8	12.03	11/11/2021	16:53

WORKLIST(Hardcopy Internal Chain)

117329

WorkList Name : CORROSIVITY M4656 w WorkList ID : 154257 Department : Wet-Chemistry Date : 11-11-2021 16:19:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/11/2021	Water	M4656-06	Corrosivity	Cool 4 deg C	PSEG03	D12	TOTE-COMP	11/11/2021	9040C
11/11/2021	Water	M4656-07	Corrosivity	Cool 4 deg C	PSEG03	D12	8076	11/11/2021	9040C
11/11/2021	Water	M4656-08	Corrosivity	Cool 4 deg C	PSEG03	D12	8025	11/11/2021	9040C
11/11/2021	Water	M4656-09	Corrosivity	Cool 4 deg C	PSEG03	D12	8089	11/11/2021	9040C

Date/Time 11/11/21 16:20
 Raw Sample Received by: JP
 Raw Sample Relinquished by: CP SM

Date/Time 11-11-21 18:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (WOC)