



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117332

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	11/12/2021	08:15
2	CAL2	1	Water	NA	NA	20.4	7.00	11/12/2021	08:16
3	CAL3	1	Water	NA	NA	20.4	10.03	11/12/2021	08:17
4	ICV	1	Water	NA	NA	20.5	7.00	11/12/2021	08:19
5	CCV1	1	Water	NA	NA	20.5	2.01	11/12/2021	08:20
6	M4660-01	1	Water	NA	NA	20.2	7.14	11/12/2021	08:22
7	M4660-01DUP	1	Water	NA	NA	20.3	7.15	11/12/2021	08:23
8	CCV2	1	Water	NA	NA	20.4	12.03	11/12/2021	08:25

117332

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4660

WorkList ID : 154261

Department : Wet-Chemistry

Date : 11-12-2021 08:05:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/11/2021	Water	M4660-01	pH	Cool 4 deg C	PSEG03	D12	FM655	11/11/2021	9040C

Date/Time 11-12-21 08:10
 Raw Sample Received by: JP/SC
 Raw Sample Relinquished by: CP SM

Date/Time 11-12-21 10:30
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP/SC