



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 11/13/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 103 °CDATE & TIME IN1: 11/12/2021 16:30 DATE & TIME OUT1: 11/13/2021 08:00OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 11/13/2021 08:30 DATE & TIME OUT2: 11/13/2021 10:00Run Number: LB117334BalanceID: WC SC-6OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB117334BL	LB117334BL	86.3148	86.3148	100	86.3148	86.3148	86.3148	0.0000	0
2	M4660-01	FM655	75.5440	75.5440	100	75.5747	75.5747	75.5747	0.0307	307
3	M4660-01DUP	FM655DUP	89.1506	89.1506	100	89.1801	89.1801	89.1801	0.0295	295
4	M4674-09	MF-WATER-COMP	81.4729	81.4729	100	81.6233	81.6233	81.6233	0.1504	1504

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

LB 117334

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS N4660 W WorkList ID : 154263 Department : Wet-Chemistry Date : 11-12-2021 08:07:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/16/2021	Water	M4660-01	N TS	Cool 4 deg C	PSEG03	D12	FM655	11/11/2021	SM2540 B
11/18/2021	Water	M4674-09	E TS	Cool 4 deg C	PSEG03	D21	MF-WATER-COMP	11/12/2021	SM2540 B

Date/Time 11.12.21 16:10
 Raw Sample Received by: JB (WCS)
 Raw Sample Relinquished by: JB (WCS)

Date/Time 11.12.21 18:30
 Raw Sample Received by: JB (WCS)
 Raw Sample Relinquished by: JB (WCS)