

LB 117337


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35514



A= 0.0000e+000

B= 2.8383e-004

C= -1.4376e-002

Rho= 0.9999097

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.012	-0.012	7	0.000	7					1
0.05	0.050	0.056	0.006	247	0.0 %	247	0				12
0.2	0.200	0.202	0.002	761	0.0 %	761					1
2.5	2.500	2.543	0.043	9010	0.0 %	9010					2
5.0	5.000	5.002	0.002	17674	0.0 %	17674					0
7.5	7.500	7.395	-0.105	26104	0.0 %	26104					1
10.0	10.000	10.065	0.065	35514	0.0 %	35514					1

LB117337 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Nov 2021 12:01:08

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	7	-	0.0000	245.1	PPB	12 Nov 2021 12:06:19	S	Std
0.05 - 1	50.05	247	-	0.0500	245.1	PPB	12 Nov 2021 12:08:36	S	Std
0.2 - 1	50.2	761	-	0.2000	245.1	PPB	12 Nov 2021 12:10:53	S	Std
2.5 - 1	50.5	9010	-	2.5000	245.1	PPB	12 Nov 2021 12:13:10	S	Std
5.0 - 1	55	17674	-	5.0000	245.1	PPB	12 Nov 2021 12:15:26	S	Std
7.5 - 1	57.5	26104	-	7.5000	245.1	PPB	12 Nov 2021 12:22:42	S	Std
10.0 - 1	510	35514	-	10.0000	245.1	PPB	12 Nov 2021 12:24:57	S	Std
ICV80 - 1	ICV80	14061	3.9765	-245.1		PPB	12 Nov 2021 12:30:20	U	SMPL
ICB80 - 1	ICB80	44	-0.0019	-245.1		PPB	12 Nov 2021 12:32:35	U	SMPL
CCV54 - 1	CCV54	17306	4.8975	-245.1		PPB	12 Nov 2021 12:34:49	U	SMPL
CCB54 - 1	CCB54	46	-0.0013	-245.1		PPB	12 Nov 2021 12:37:05	U	SMPL
CRA - 1	CRA	234	0.0520	-245.1		PPB	12 Nov 2021 12:39:19	U	SMPL
HighStd - 1	HighStd	35664	10.1081	-245.1		PPB	12 Nov 2021 12:41:34	U	SMPL
ChkStd - 1	ChkStd	24564	6.9576	-245.1		PPB	12 Nov 2021 12:43:50	U	SMPL
PB140732BL - 1	PBW	33	-0.0050	-245.1		PPB	12 Nov 2021 12:46:07	U	SMPL
PB140732BS - 1	LCSW	12938	3.6578	-245.1		PPB	12 Nov 2021 12:48:24	U	SMPL
M4068-10 - 1	LOD-MDL-WATER-04-QT4-2021	549	0.1414	-245.1		PPB	12 Nov 2021 12:50:40	U	SMPL
M4068-12 - 1	MDL-WATER-06-QT4-2021	234	0.0520	-245.1		PPB	12 Nov 2021 12:52:56	U	SMPL
CCV55 - 1	CCV55	17276	4.8890	-245.1		PPB	12 Nov 2021 12:57:21	U	SMPL
CCB55 - 1	CCB55	11	-0.0113	-245.1		PPB	12 Nov 2021 12:59:38	U	SMPL