



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 11/15/2021

OVENTEMP IN Celsius(°C): 108
Time IN: 14:00
In Date: 11/12/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/13/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117338

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4632-01	GB8P4	1	1.08	8.82	9.9	8.41	83.1	
M4632-02	GB8P4D	2	1.08	8.82	9.9	8.41	83.1	
M4632-03	GB8P4S	3	1.08	8.82	9.9	8.41	83.1	
M4632-04	GB8P5	4	1.08	8.82	9.9	7.41	71.8	
M4632-05	GB8P6	5	1.08	8.72	9.8	7.44	72.9	
M4632-06	GB8P7	6	1.09	8.71	9.8	7.89	78.1	
M4632-07	GB8P8	7	1.08	8.72	9.8	8.00	79.4	
M4632-08	GB8P9	8	1.08	8.62	9.7	8.04	80.7	
M4632-09	GB8Q0	9	1.08	8.87	9.95	7.74	75.1	
M4632-10	GB8Q1	10	1.08	8.82	9.9	7.96	78.0	
M4632-11	GB8Q2	11	1.08	8.77	9.85	6.18	58.2	
M4632-12	GB8Q3	12	1.08	8.62	9.7	7.92	79.4	
M4632-13	GB8Q4	13	1.08	8.72	9.8	6.71	64.6	
M4632-14	GB8Q5	14	1.08	8.72	9.8	7.26	70.9	
M4632-15	GB8Q6	15	1.08	8.89	9.97	7.51	72.3	
M4632-16	GB8Q7	16	1.08	8.43	9.51	8.03	82.4	
M4632-17	GB8Q8	17	1.08	8.77	9.85	8.49	84.5	
M4632-18	GB8Q9	18	1.08	8.89	9.97	6.67	62.9	
M4632-19	GB8R0	19	1.08	8.90	9.98	7.41	71.1	
M4632-20	GB8R1	20	1.08	8.57	9.65	6.82	67.0	
M4632-21	GB8R2	21	1.08	8.53	9.61	6.5	63.5	
M4632-22	GB8R3	22	1.08	8.72	9.8	7.49	73.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$