


LB117342

7471B

INSTRUMENT ID: CV1

μ Abs.:

54169



Linear

A= 0.0000e+000

B= 1.8485e-004

C= 6.1951e-004

Rho= 0.9999778

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	0.008	0.008	39	0.000	39					3
0.2	0.200	0.207	0.007	1118	0.0 %	1118					1
2.5	2.500	2.468	-0.032	13346	0.0 %	13346					0
5.0	5.000	5.035	0.035	27236	0.0 %	27236					0
7.5	7.500	7.468	-0.032	40396	0.0 %	40396					0
10.0	10.000	10.014	0.014	54169	0.0 %	54169					0

LB117348 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Nov 2021 14:41:35

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	39	-	0.00007471B	PPB	12 Nov 2021 14:42:40	S	Std
0.2 - 1	50.2	1118	-	0.20007471B	PPB	12 Nov 2021 14:44:57	S	Std
2.5 - 1	52.5	13346	-	2.50007471B	PPB	12 Nov 2021 14:47:14	S	Std
5.0 - 1	55	27236	-	5.00007471B	PPB	12 Nov 2021 14:49:30	S	Std
7.5 - 1	57.5	40396	-	7.50007471B	PPB	12 Nov 2021 14:51:47	S	Std
10.0 - 1	510	54169	-	10.00007471B	PPB	12 Nov 2021 14:54:03	S	Std
ICV81 - 1	ICV81	20758	3.8378	-7471B	PPB	12 Nov 2021 14:56:55	U	SMPL
ICB81 - 1	ICB81	-58	-0.0101	-7471B	PPB	12 Nov 2021 14:59:10	U	SMPL
CCV56 - 1	CCV56	26721	4.9401	-7471B	PPB	12 Nov 2021 15:01:24	U	SMPL
CCB56 - 1	CCB56	-70	-0.0123	-7471B	PPB	12 Nov 2021 15:03:39	U	SMPL
CRA - 1	CRA	1087	0.2016	-7471B	PPB	12 Nov 2021 15:05:55	U	SMPL
HighStd - 1	HighStd	52431	9.6927	-7471B	PPB	12 Nov 2021 15:08:09	U	SMPL
ChkStd - 1	ChkStd	36691	6.7831	-7471B	PPB	12 Nov 2021 15:10:25	U	SMPL
PB140733BL - 1	PBS	-98	-0.0175	-7471B	PPB	12 Nov 2021 15:12:42	U	SMPL
PB140733BS - 1	LCSS	19697	3.6417	-7471B	PPB	12 Nov 2021 15:15:00	U	SMPL
M4639-01 - 1	TT-1/2-3	3214	0.5947	-7471B	PPB	12 Nov 2021 15:22:05	U	SMPL
M4639-01DUP - 1	TT-1/2-3DUP	3257	0.6027	-7471B	PPB	12 Nov 2021 15:24:21	U	SMPL
M4639-01MS - 1	TT-1/2-3MS	22990	4.2504	-7471B	PPB	12 Nov 2021 15:26:38	U	SMPL
M4639-01MSD - 1	TT-1/2-3MSD	22051	4.0768	-7471B	PPB	12 Nov 2021 15:28:55	U	SMPL
M4639-02 - 1	TT-3/4-5	1771	0.3280	-7471B	PPB	12 Nov 2021 15:31:10	U	SMPL
CCV57 - 1	CCV57	26317	4.8654	-7471B	PPB	12 Nov 2021 15:33:26	U	SMPL
CCB57 - 1	CCB57	-42	-0.0071	-7471B	PPB	12 Nov 2021 15:35:41	U	SMPL
M4639-03 - 1	TT-4/1-2	2734	0.5060	-7471B	PPB	12 Nov 2021 15:37:56	U	SMPL
M4639-04 - 1	TT-5/5.5-6.5	2047	0.3790	-7471B	PPB	12 Nov 2021 15:40:11	U	SMPL
M4639-05 - 1	TT-6/4-5	2137	0.3957	-7471B	PPB	12 Nov 2021 15:42:27	U	SMPL
M4639-06 - 1	TT-7/1-2	2027	0.3753	-7471B	PPB	12 Nov 2021 15:44:42	U	SMPL
M4639-07 - 1	TT-8/0.5-1.5	1424	0.2639	-7471B	PPB	12 Nov 2021 15:46:58	U	SMPL
M4639-08 - 1	TT-9/4-5	2208	0.4088	-7471B	PPB	12 Nov 2021 15:49:15	U	SMPL
M4639-09 - 1	TT-10/0-1	2860	0.5293	-7471B	PPB	12 Nov 2021 15:51:31	U	SMPL
M4639-10 - 1	TT-11/2-3	2473	0.4578	-7471B	PPB	12 Nov 2021 15:53:48	U	SMPL
M4640-01 - 1	WC-2-AS	19429	3.5921	-7471B	PPB	12 Nov 2021 15:56:05	U	SMPL
M4651-01 - 1	HD-02-111121	1579	0.2925	-7471B	PPB	12 Nov 2021 15:58:20	U	SMPL
CCV58 - 1	CCV58	26632	4.9236	-7471B	PPB	12 Nov 2021 16:00:36	U	SMPL
CCB58 - 1	CCB58	-19	-0.0029	-7471B	PPB	12 Nov 2021 16:02:51	U	SMPL
M4653-01 - 1	CONCRETE-GRAB	7449	1.3776	-7471B	PPB	12 Nov 2021 16:05:07	U	SMPL
M4654-01 - 1	PARKINGLOT-BOTTOM-U-SP	395	0.0736	-7471B	PPB	12 Nov 2021 16:07:23	U	SMPL
M4654-04 - 1	PARKINGLOT-TOP-U-SP	1332	0.2468	-7471B	PPB	12 Nov 2021 16:09:38	U	SMPL
M4657-01 - 1	VWE-B31-111121-0-10	3567	0.6600	-7471B	PPB	12 Nov 2021 16:11:54	U	SMPL
M4657-02 - 1	VWE-B31-111121-10-15	460	0.0857	-7471B	PPB	12 Nov 2021 16:14:10	U	SMPL
M4659-01 - 1	HB-2-(6-6.5)-11-10-21	1368	0.2535	-7471B	PPB	12 Nov 2021 16:16:26	U	SMPL
M4659-02 - 1	HB-1-(7.5-8)-11-11-21	1981	0.3668	-7471B	PPB	12 Nov 2021 16:18:43	U	SMPL
M4639-01LX5 - 1		397	0.0740	-7471B	PPB	12 Nov 2021 16:20:59	U	SMPL
M4639-01A - 1		21521	3.9789	-7471B	PPB	12 Nov 2021 16:35:13	U	SMPL
CCV59 - 1	CCV59	24659	4.5589	-7471B	PPB	12 Nov 2021 16:37:28	U	SMPL
CCB59 - 1	CCB59	-54	-0.0094	-7471B	PPB	12 Nov 2021 16:39:45	U	SMPL