



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117355

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	11/12/2021	16:40
2	CAL2	1	Water	NA	NA	20.6	7.00	11/12/2021	16:41
3	CAL3	1	Water	NA	NA	20.6	10.03	11/12/2021	16:42
4	ICV	1	Water	NA	NA	20.7	7.01	11/12/2021	16:45
5	CCV1	1	Water	NA	NA	20.7	2.01	11/12/2021	16:50
6	M4674-09	1	Water	NA	NA	20.3	11.25	11/12/2021	17:10
7	M4674-09DUP	1	Water	NA	NA	20.4	11.26	11/12/2021	17:11
8	CCV2	1	Water	NA	NA	20.7	12.03	11/12/2021	17:12

WORKLIST(Hardcopy Internal Chain)

VB 117355

WorkList Name : PH M4670 W WorkList ID : 154303 Department : Wet-Chemistry Date : 11-12-2021 16:34:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/12/2021	Water	M4674-09	pH	Cool 4 deg C	PSEG03	D21	MF-WATER-COMP	11/12/2021	9040C

Date/Time 11-12-21 16:37
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 11-12-21 18:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]