



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117356

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	11/12/2021	16:40
2	CAL2	1	Water	NA	NA	20.6	7.00	11/12/2021	16:41
3	CAL3	1	Water	NA	NA	20.6	10.03	11/12/2021	16:42
4	ICV	1	Water	NA	NA	20.7	7.01	11/12/2021	16:45
5	CCV1	1	Water	NA	NA	20.7	2.01	11/12/2021	16:50
6	M4670-01	1	Solid	20.03	20	21.6	6.60	11/12/2021	17:00
7	M4684-01	1	Solid	20.02	20	20.9	9.12	11/12/2021	17:01
8	M4684-01DUP	1	Solid	20.03	20	21.1	9.13	11/12/2021	17:02
9	CCV2	1	Water	NA	NA	20.8	12.03	11/12/2021	17:05

117356

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4674 WorkList ID : 154302 Department : Wet-Chemistry Date : 11-12-2021 16:33:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/12/2021	Solid	M4670-01	pH	Cool 4 deg C	PSEG03	D21	NWB-1707	11/12/2021	9045D
11/12/2021	Solid	M4684-01	pH	Cool 4 deg C	PSEG05	D21	AU-5-111221	11/12/2021	9045D

Date/Time 11.12.21 16:35
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11.12.21 18:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]