


LB117363

7470A

INSTRUMENT ID: CV1

 μ Abs.:

37682



Linear

A= 0.0000e+000

B= 2.6442e-004

C= -3.3464e-003

Rho= 0.9998994

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.004	0.004	26	0.000	26					
0.2	0.200	0.176	-0.024	679	0.0 %	679					-12
2.5	2.500	2.457	-0.043	9306	0.0 %	9306					-2
5.0	5.000	5.111	0.111	19343	0.0 %	19343					2
7.5	7.500	7.491	-0.009	28342	0.0 %	28342					0
10.0	10.000	9.961	-0.039	37682	0.0 %	37682					0

LB117363

Method: 7470A

Operator: Admin

Date of Analysis: 15 Nov 2021 09:02:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	26	-	0.0000	7470A	PPB	15 Nov 2021 09:50:15	S	Std
0.2 - 1	SP-2	679	-	0.2000	7470A	PPB	15 Nov 2021 09:52:34	S	Std
2.5 - 1	SP-5	9306	-	2.5000	7470A	PPB	15 Nov 2021 09:54:52	S	Std
5.0 - 1	SP-5	19343	-	5.0000	7470A	PPB	15 Nov 2021 09:57:09	S	Std
7.5 - 1	SP-5	28342	-	7.5000	7470A	PPB	15 Nov 2021 09:59:25	S	Std
10.0 - 1	SP-10	37682	-	10.0000	7470A	PPB	15 Nov 2021 10:01:41	S	Std
ICV82 - 1	ICV82	13990	3.6959	-	7470A	PPB	15 Nov 2021 10:04:28	U	SMPL
ICB82 - 1	ICB82	-50	-0.0166	-	7470A	PPB	15 Nov 2021 10:06:42	U	SMPL
CCV60 - 1	CCV60	18746	4.9535	-	7470A	PPB	15 Nov 2021 10:08:57	U	SMPL
CCB60 - 1	CCB60	-51	-0.0168	-	7470A	PPB	15 Nov 2021 10:11:12	U	SMPL
CRA - 1	CRA	730	0.1897	-	7470A	PPB	15 Nov 2021 10:13:27	U	SMPL
HighStd - 1	HighStd	37929	10.0259	-	7470A	PPB	15 Nov 2021 10:15:42	U	SMPL
ChkStd - 1	ChkStd	26800	7.0832	-	7470A	PPB	15 Nov 2021 10:17:57	U	SMPL
PB140734BL - 1	PBW	-107	-0.0316	-	7470A	PPB	15 Nov 2021 10:20:14	U	SMPL
PB140734BS - 1	LCSW	13301	3.5137	-	7470A	PPB	15 Nov 2021 10:22:31	U	SMPL
M4640-02 - 1	WC-2-AS	-28	-0.0108	-	7470A	PPB	15 Nov 2021 10:24:47	U	SMPL
M4640-02DUP - 1	WC-2-ASDUP	19	0.0017	-	7470A	PPB	15 Nov 2021 10:27:04	U	SMPL
M4640-02MS - 1	WC-2-ASMS	13673	3.6121	-	7470A	PPB	15 Nov 2021 10:29:22	U	SMPL
M4640-02MSD - 1	WC-2-ASMSD	13962	3.6885	-	7470A	PPB	15 Nov 2021 10:31:39	U	SMPL
M4644-02 - 1	C-5	-35	-0.0126	-	7470A	PPB	15 Nov 2021 10:33:55	U	SMPL
CCV61 - 1	CCV61	18844	4.9794	-	7470A	PPB	15 Nov 2021 10:36:10	U	SMPL
CCB61 - 1	CCB61	-39	-0.0137	-	7470A	PPB	15 Nov 2021 10:38:26	U	SMPL
M4656-01 - 1	D-COMP1	20	0.0019	-	7470A	PPB	15 Nov 2021 10:40:41	U	SMPL
M4656-02 - 1	D-COMP2	22	0.0025	-	7470A	PPB	15 Nov 2021 10:42:57	U	SMPL
M4656-03 - 1	D-COMP3	39	0.0070	-	7470A	PPB	15 Nov 2021 10:45:12	U	SMPL
M4656-04 - 1	D-COMP4	51	0.0101	-	7470A	PPB	15 Nov 2021 10:47:28	U	SMPL
M4656-05 - 1	7921	719	0.1868	-	7470A	PPB	15 Nov 2021 10:49:44	U	SMPL
PB140735BL - 1	PBW	-203	-0.0570	-	7470A	PPB	15 Nov 2021 10:52:00	U	SMPL
PB140735BS - 1	LCSW	13513	3.5698	-	7470A	PPB	15 Nov 2021 10:54:17	U	SMPL
M4656-06 - 1	TOTE-COMP	-31	-0.0115	-	7470A	PPB	15 Nov 2021 10:56:34	U	SMPL
M4656-07 - 1	8076	102	0.0236	-	7470A	PPB	15 Nov 2021 10:58:51	U	SMPL
M4656-08 - 1	8025	31	0.0049	-	7470A	PPB	15 Nov 2021 11:01:07	U	SMPL
CCV62 - 1	CCV62	19069	5.0389	-	7470A	PPB	15 Nov 2021 11:03:22	U	SMPL
CCB62 - 1	CCB62	-46	-0.0155	-	7470A	PPB	15 Nov 2021 11:05:38	U	SMPL
M4656-08DUP - 1	8025DUP	19	0.0017	-	7470A	PPB	15 Nov 2021 11:07:54	U	SMPL
M4656-08MS - 1	8025MS	13462	3.5563	-	7470A	PPB	15 Nov 2021 11:10:09	U	SMPL
M4656-08MSD - 1	8025MSD	14244	3.7631	-	7470A	PPB	15 Nov 2021 11:12:25	U	SMPL
M4656-09 - 1	8089	59	0.0123	-	7470A	PPB	15 Nov 2021 11:14:40	U	SMPL
PB140736BL - 1	PBW	44	0.0083	-	7470A	PPB	15 Nov 2021 11:16:56	U	SMPL
PB140736BS - 1	LCSW	14535	3.8400	-	7470A	PPB	15 Nov 2021 11:19:13	U	SMPL
M4660-01 - 1	FM655	29	0.0043	-	7470A	PPB	15 Nov 2021 11:21:29	U	SMPL
M4660-01DUP - 1	FM655DUP	40	0.0072	-	7470A	PPB	15 Nov 2021 11:23:46	U	SMPL
M4660-01MS - 1	FM655MS	12774	3.3744	-	7470A	PPB	15 Nov 2021 11:26:03	U	SMPL
M4660-01MSD - 1	FM655MSD	13089	3.4577	-	7470A	PPB	15 Nov 2021 11:28:19	U	SMPL
CCV63 - 1	CCV63	19509	5.1553	-	7470A	PPB	15 Nov 2021 11:30:34	U	SMPL
CCB63 - 1	CCB63	-55	-0.0179	-	7470A	PPB	15 Nov 2021 11:32:50	U	SMPL
PB140734TB - 1	PB140713TB	20	0.0019	-	7470A	PPB	15 Nov 2021 11:35:06	U	SMPL
PB140735TB - 1	PB140719TB	28	0.0041	-	7470A	PPB	15 Nov 2021 11:37:21	U	SMPL
M4640-02LX5 - 1		21	0.0022	-	7470A	PPB	15 Nov 2021 11:39:37	U	SMPL
M4640-02A - 1		14076	3.7187	-	7470A	PPB	15 Nov 2021 11:41:53	U	SMPL
M4656-08LX5 - 1		-13	-0.0068	-	7470A	PPB	15 Nov 2021 11:44:09	U	SMPL
M4656-08A - 1		13882	3.6674	-	7470A	PPB	15 Nov 2021 11:46:25	U	SMPL
M4660-01LX5 - 1		0	-0.0033	-	7470A	PPB	15 Nov 2021 11:48:41	U	SMPL
M4660-01A - 1		13834	3.6547	-	7470A	PPB	15 Nov 2021 11:50:58	U	SMPL
CCV64 - 1	CCV64	18748	4.9540	-	7470A	PPB	15 Nov 2021 11:53:15	U	SMPL
CCB64 - 1	CCB64	-18	-0.0081	-	7470A	PPB	15 Nov 2021 11:55:31	U	SMPL