



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117366

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/13/2021	09:30
2	CAL2	1	Water	NA	NA	20.3	7.00	11/13/2021	09:31
3	CAL3	1	Water	NA	NA	20.3	10.03	11/13/2021	09:32
4	ICV	1	Water	NA	NA	20.4	6.99	11/13/2021	09:35
5	CCV1	1	Water	NA	NA	20.3	2.01	11/13/2021	09:36
6	M4693-02	1	Solid	20.03	20	20.2	8.69	11/13/2021	09:40
7	M4693-02DUP	1	Solid	20.04	20	20.3	8.70	11/13/2021	09:41
8	M4693-04	1	Solid	20.05	20	20.3	9.10	11/13/2021	09:45
9	CCV2	1	Water	NA	NA	20.4	12.03	11/13/2021	09:47

WORKLIST(Hardcopy Internal Chain)

117366
VB

WorkList Name : CORROSIONITY M4670 WorkList ID : 154312 Department : Wet-Chemistry Date : 11-13-2021 09:21:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/12/2021	Solid	M4693-02	Corrosivity	Cool 4 deg C	WALS01	D21	UT-COMP-01	11/12/2021	9045D
11/12/2021	Solid	M4693-04	Corrosivity	Cool 4 deg C	WALS01	D21	UT-COMP-02	11/12/2021	9045D

Date/Time 11-13-21 9:25
 Raw Sample Received by: JB LWC
 Raw Sample Relinquished by: CP SM

Date/Time 11-13-21 11:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JB LWC