

LB 117383


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

38872



A= 0.0000e+000

B= 2.5883e-004

C= -4.6657e-002

Rho= 0.9994907

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	173	0.000	173					-
0.2	0.200	0.168	-0.032	831	0.0 %	831					-16
2.5	2.500	2.459	-0.041	9680	0.0 %	9680					-2
5.0	5.000	5.228	0.228	20379	0.0 %	20379					5
7.5	7.500	7.332	-0.168	28507	0.0 %	28507					-2
10.0	10.000	10.015	0.015	38872	0.0 %	38872					0

LB117383 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 15 Nov 2021 15:10:24					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	173	-	0.0000	7470A	PPB	15 Nov 2021 15:11:45	S	Std
0.2 - 1	50.2	831	-	0.2000	7470A	PPB	15 Nov 2021 15:14:02	S	Std
2.5 - 1	52.5	9680	-	2.5000	7470A	PPB	15 Nov 2021 15:16:19	S	Std
5.0 - 1	55	20379	-	5.0000	7470A	PPB	15 Nov 2021 15:18:36	S	Std
7.5 - 1	57.5	28507	-	7.5000	7470A	PPB	15 Nov 2021 15:20:52	S	Std
10.0 - 1	510	38872	-	10.0000	7470A	PPB	15 Nov 2021 15:25:41	S	Std
ICV83 - 1	ICV83	16689	4.2730	-	7470A	PPB	15 Nov 2021 15:30:47	U	SMPL
ICB83 - 1	ICB83	40	-0.0363	-	7470A	PPB	15 Nov 2021 15:33:01	U	SMPL
CCV65 - 1	CCV65	19452	4.9881	-	7470A	PPB	15 Nov 2021 15:35:16	U	SMPL
CCB65 - 1	CCB65	52	-0.0332	-	7470A	PPB	15 Nov 2021 15:37:31	U	SMPL
CRA - 1	CRA	800	0.1604	-	7470A	PPB	15 Nov 2021 15:55:34	U	SMPL
HighStd - 1	HighStd	39327	10.1324	-	7470A	PPB	15 Nov 2021 15:57:48	U	SMPL
ChkStd - 1	ChkStd	27994	7.1991	-	7470A	PPB	15 Nov 2021 16:00:04	U	SMPL
PB140788BL - 1	PBW	37	-0.0371	-	7470A	PPB	15 Nov 2021 16:02:20	U	SMPL
PB140788BS - 1	LCSW	15305	3.9148	-	7470A	PPB	15 Nov 2021 16:04:36	U	SMPL
M4603-02 - 1	ISCO-SOIL-IDW-11082021	-25	-0.0531	-	7470A	PPB	15 Nov 2021 16:06:53	U	SMPL
M4670-01 - 1	NWB-1707	103	-0.0200	-	7470A	PPB	15 Nov 2021 16:09:09	U	SMPL
M4693-02 - 1	UT-COMP-01	117	-0.0164	-	7470A	PPB	15 Nov 2021 16:11:26	U	SMPL
M4693-04 - 1	UT-COMP-02	124	-0.0146	-	7470A	PPB	15 Nov 2021 16:13:43	U	SMPL
M4693-04DUP - 1	UT-COMP-02DUP	28	-0.0394	-	7470A	PPB	15 Nov 2021 16:15:58	U	SMPL
CCV66 - 1	CCV66	20677	5.3052	-	7470A	PPB	15 Nov 2021 16:18:13	U	SMPL
CCB66 - 1	CCB66	42	-0.0358	-	7470A	PPB	15 Nov 2021 16:20:29	U	SMPL
M4693-04MS - 1	UT-COMP-02MS	13449	3.4344	-	7470A	PPB	15 Nov 2021 16:22:44	U	SMPL
M4693-04MSD - 1	UT-COMP-02MSD	13609	3.4758	-	7470A	PPB	15 Nov 2021 16:25:00	U	SMPL
PB140795BL - 1	PBW	62	-0.0306	-	7470A	PPB	15 Nov 2021 16:27:15	U	SMPL
PB140795BS - 1	LCSW	14104	3.6039	-	7470A	PPB	15 Nov 2021 16:29:31	U	SMPL
M4674-09 - 1	MF-WATER-COMP	604	0.1097	-	7470A	PPB	15 Nov 2021 16:34:34	U	SMPL
M4674-09DUP - 1	MF-WATER-COMPDUP	717	0.1389	-	7470A	PPB	15 Nov 2021 16:39:42	U	SMPL
M4674-09MS - 1	MF-WATER-COMPMS	15884	4.0646	-	7470A	PPB	15 Nov 2021 16:41:58	U	SMPL
M4674-09MSD - 1	MF-WATER-COMPMSD	18101	4.6385	-	7470A	PPB	15 Nov 2021 16:44:15	U	SMPL
M4687-04 - 1	FIELD-BLANK	72	-0.0280	-	7470A	PPB	15 Nov 2021 16:46:32	U	SMPL
CCV67 - 1	CCV67	20704	5.3122	-	7470A	PPB	15 Nov 2021 16:48:48	U	SMPL
CCB67 - 1	CCB67	75	-0.0272	-	7470A	PPB	15 Nov 2021 16:51:03	U	SMPL
PB140795TB - 1	PB140770TB	106	-0.0192	-	7470A	PPB	15 Nov 2021 16:53:19	U	SMPL
M4693-04LX5 - 1		56	-0.0322	-	7470A	PPB	15 Nov 2021 16:55:34	U	SMPL
M4693-04A - 1		15237	3.8972	-	7470A	PPB	15 Nov 2021 16:57:50	U	SMPL
M4674-09LX5 - 1		396	0.0558	-	7470A	PPB	15 Nov 2021 17:00:05	U	SMPL
M4674-09A - 1		16659	4.2652	-	7470A	PPB	15 Nov 2021 17:02:21	U	SMPL
CCV68 - 1	CCV68	20385	5.2296	-	7470A	PPB	15 Nov 2021 17:04:36	U	SMPL
CCB68 - 1	CCB68	71	-0.0283	-	7470A	PPB	15 Nov 2021 17:06:53	U	SMPL