



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117388

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	11/15/2021	13:58
2	M4656-05		1	8	63.00	145.4	145.4	11/15/2021	17:20
3	M4656-05DUP		1	8	64.00	147.2	147.2	11/15/2021	17:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117388

WorkList Name : fp111521-1 WorkList ID : 154356 Department : Wet-Chemistry Date : 11-15-2021 15:00:54

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/18/2021	Solid	M4656-05	B		Cool 4 deg C	PSEG03	D12	7921	11/11/2021	1010A

Date/Time 11/15/21 15:15
 Raw Sample Received by: FS (WC)
 Raw Sample Relinquished by: FS (WC)

Date/Time 11/15/21 17:50
 Raw Sample Received by: FS (WC)
 Raw Sample Relinquished by: FS (WC)