


LB117391

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.029	0.029	191	0.000	191					5
0.2	0.200	0.210	0.010	1150	0.0 %	1150					-3
2.5	2.500	2.437	-0.063	12945	0.0 %	12945					0
5.0	5.000	5.001	0.001	26531	0.0 %	26531					0
7.5	7.500	7.533	0.033	39941	0.0 %	39941					0
10.0	10.000	9.990	-0.010	52960	0.0 %	52960					0

LB117391 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Nov 2021 09:03:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	191	-	0.00007471B	PPB		16 Nov 2021 09:26:38	S	Std
0.2 - 1	50.2	1150	-	0.20007471B	PPB		16 Nov 2021 09:28:56	S	Std
2.5 - 1	52.5	12945	-	2.50007471B	PPB		16 Nov 2021 09:31:13	S	Std
5.0 - 1	55	26531	-	5.00007471B	PPB		16 Nov 2021 09:33:29	S	Std
7.5 - 1	57.5	39941	-	7.50007471B	PPB		16 Nov 2021 09:35:46	S	Std
10.0 - 1	510	52960	-	10.00007471B	PPB		16 Nov 2021 09:38:03	S	Std
ICV84 - 1	ICV84	20189	3.8041	-7471B	PPB		16 Nov 2021 09:41:00	U	SMPL
ICB84 - 1	ICB84	269	0.0440	-7471B	PPB		16 Nov 2021 09:43:15	U	SMPL
CCV69 - 1	CCV69	26864	5.0641	-7471B	PPB		16 Nov 2021 09:45:30	U	SMPL
CCB69 - 1	CCB69	6	-0.0057	-7471B	PPB		16 Nov 2021 09:47:45	U	SMPL
CRA - 1	CRA	1176	0.2152	-7471B	PPB		16 Nov 2021 09:50:01	U	SMPL
HighStd - 1	HighStd	52829	9.9653	-7471B	PPB		16 Nov 2021 09:52:16	U	SMPL
ChkStd - 1	ChkStd	36549	6.8922	-7471B	PPB		16 Nov 2021 09:54:31	U	SMPL
PB140793BL - 1	PBS	66	0.0057	-7471B	PPB		16 Nov 2021 09:56:48	U	SMPL
PB140793BS - 1	LCSS	20191	3.8045	-7471B	PPB		16 Nov 2021 09:59:04	U	SMPL
M4665-01 - 1	ASPHALT	5524	1.0359	-7471B	PPB		16 Nov 2021 10:01:21	U	SMPL
M4665-01DUP - 1	ASPHALTDUP	5308	0.9951	-7471B	PPB		16 Nov 2021 10:03:37	U	SMPL
M4665-01MS - 1	ASPHALTMS	23776	4.4812	-7471B	PPB		16 Nov 2021 10:08:34	U	SMPL
M4665-01MSD - 1	ASPHALTMSD	24582	4.6333	-7471B	PPB		16 Nov 2021 10:10:51	U	SMPL
M4666-01 - 1	CONCRETE	7732	1.4527	-7471B	PPB		16 Nov 2021 10:13:06	U	SMPL
CCV70 - 1	CCV70	26971	5.0843	-7471B	PPB		16 Nov 2021 10:15:22	U	SMPL
CCB70 - 1	CCB70	279	0.0459	-7471B	PPB		16 Nov 2021 10:17:37	U	SMPL
M4667-01 - 1	ASPHALT-CHUNK	518	0.0910	-7471B	PPB		16 Nov 2021 10:19:53	U	SMPL
M4673-01 - 1	RO-72-11933	4837	0.9062	-7471B	PPB		16 Nov 2021 10:22:08	U	SMPL
M4674-01 - 1	MF-SOIL-P1	1478	0.2722	-7471B	PPB		16 Nov 2021 10:24:24	U	SMPL
M4674-03 - 1	MF-SOIL-P2	1011	0.1840	-7471B	PPB		16 Nov 2021 10:26:40	U	SMPL
M4674-05 - 1	MF-SOIL-P3	800	0.1442	-7471B	PPB		16 Nov 2021 10:28:56	U	SMPL
M4674-07 - 1	MF-SOIL-P4	884	0.1601	-7471B	PPB		16 Nov 2021 10:31:12	U	SMPL
M4679-01 - 1	LCP1-SW-63	17036	3.2089	-7471B	PPB		16 Nov 2021 10:33:29	U	SMPL
M4679-02 - 1	LCP1-SW-61	19315	3.6391	-7471B	PPB		16 Nov 2021 10:35:45	U	SMPL
M4679-03 - 1	LCP1-SED-110	12752	2.4003	-7471B	PPB		16 Nov 2021 10:38:02	U	SMPL
M4680-01 - 1	COMP-1	14806	2.7880	-7471B	PPB		16 Nov 2021 10:40:18	U	SMPL
CCV71 - 1	CCV71	27276	5.1419	-7471B	PPB		16 Nov 2021 10:42:34	U	SMPL
CCB71 - 1	CCB71	295	0.0489	-7471B	PPB		16 Nov 2021 10:44:49	U	SMPL
M4680-02 - 1	COMP-2	6786	1.2741	-7471B	PPB		16 Nov 2021 10:47:05	U	SMPL
M4680-03 - 1	COMP-3	10825	2.0365	-7471B	PPB		16 Nov 2021 10:49:20	U	SMPL
PB140794BL - 1	PBS	228	0.0362	-7471B	PPB		16 Nov 2021 10:51:36	U	SMPL
PB140794BS - 1	LCSS	18664	3.5162	-7471B	PPB		16 Nov 2021 10:53:51	U	SMPL
M4681-01 - 1	COMP-1	1244	0.2280	-7471B	PPB		16 Nov 2021 10:56:07	U	SMPL
M4681-01DUP - 1	COMP-1DUP	1308	0.2401	-7471B	PPB		16 Nov 2021 10:58:23	U	SMPL
M4681-01MS - 1	COMP-1MS	20676	3.8960	-7471B	PPB		16 Nov 2021 11:00:39	U	SMPL
M4681-01MSD - 1	COMP-1MSD	20425	3.8487	-7471B	PPB		16 Nov 2021 11:02:56	U	SMPL
M4681-02 - 1	COMP-2	1279	0.2346	-7471B	PPB		16 Nov 2021 11:05:13	U	SMPL
M4681-03 - 1	COMP-3	1636	0.3020	-7471B	PPB		16 Nov 2021 11:07:29	U	SMPL
CCV72 - 1	CCV72	27147	5.1175	-7471B	PPB		16 Nov 2021 11:09:44	U	SMPL
CCB72 - 1	CCB72	11	-0.0047	-7471B	PPB		16 Nov 2021 11:16:53	U	SMPL
M4683-01 - 1	NB-07-111221	1235	0.2263	-7471B	PPB		16 Nov 2021 11:19:08	U	SMPL
M4684-01 - 1	AU-5-111221	1578	0.2911	-7471B	PPB		16 Nov 2021 11:21:23	U	SMPL
M4685-01 - 1	OK-01-111221	3077	0.5740	-7471B	PPB		16 Nov 2021 11:23:39	U	SMPL
M4687-01 - 1	VWE-B35-111221-0-10	5296	0.9929	-7471B	PPB		16 Nov 2021 11:25:55	U	SMPL
M4693-01 - 1	UT-COMP-01	110661	20.8817	-7471B	PPB		16 Nov 2021 11:28:10	U	SMPL
M4693-03 - 1	UT-COMP-02	25161	4.7426	-7471B	PPB		16 Nov 2021 11:30:26	U	SMPL
M4665-01LX5 - 1		406	0.0698	-7471B	PPB		16 Nov 2021 11:32:45	U	SMPL
M4665-01A - 1		27136	5.1154	-7471B	PPB		16 Nov 2021 11:35:02	U	SMPL
M4681-01LX5 - 1		139	0.0194	-7471B	PPB		16 Nov 2021 11:37:19	U	SMPL
M4681-01A - 1		22852	4.3068	-7471B	PPB		16 Nov 2021 11:39:35	U	SMPL
CCV73 - 1	CCV73	26967	5.0835	-7471B	PPB		16 Nov 2021 11:41:50	U	SMPL
CCB73 - 1	CCB73	-36	-0.0136	-7471B	PPB		16 Nov 2021 11:44:06	U	SMPL
M4693-01DLX5 - 1		22764	4.2902	-7471B	PPB		16 Nov 2021 11:53:47	U	SMPL
CCV74 - 1	CCV74	26736	5.0399	-7471B	PPB		16 Nov 2021 11:56:02	U	SMPL
CCB74 - 1	CCB74	-40	-0.0143	-7471B	PPB		16 Nov 2021 11:58:18	U	SMPL