

LB119394


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53056



A= 0.0000e+000

B= 1.8828e-004

C= -3.1130e-002

Rho= 0.9999716

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.002	-0.002	156	0.000	156					-12
0.2	0.200	0.176	-0.024	1099	0.0 %	1099					0
2.5	2.500	2.500	-0.000	13442	0.0 %	13442					1
5.0	5.000	5.037	0.037	26916	0.0 %	26916					0
7.5	7.500	7.531	0.031	40165	0.0 %	40165					0
10.0	10.000	9.958	-0.042	53056	0.0 %	53056					0

LB117394 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Nov 2021 14:11:31

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SO	156	-	0.00007471B	PPB	16 Nov 2021 14:13:33	S	Std
0.2 - 1	SO.2	1099	-	0.20007471B	PPB	16 Nov 2021 14:15:50	S	Std
2.5 - 1	SO.5	13442	-	2.50007471B	PPB	16 Nov 2021 14:18:07	S	Std
5.0 - 1	SB	26916	-	5.00007471B	PPB	16 Nov 2021 14:20:24	S	Std
7.5 - 1	SB.5	40165	-	7.50007471B	PPB	16 Nov 2021 14:22:41	S	Std
10.0 - 1	SB.1	53056	-	10.00007471B	PPB	16 Nov 2021 14:24:57	S	Std
ICV85 - 1	ICV85	19723	3.6824	-7471B	PPB	16 Nov 2021 14:27:39	U	SMPL
ICB85 - 1	ICB85	3	-0.0306	-7471B	PPB	16 Nov 2021 14:29:53	U	SMPL
CCV75 - 1	CCV75	26285	4.9179	-7471B	PPB	16 Nov 2021 14:32:09	U	SMPL
CCB75 - 1	CCB75	-6	-0.0323	-7471B	PPB	16 Nov 2021 14:34:24	U	SMPL
CRA - 1	CRA	1070	0.1703	-7471B	PPB	16 Nov 2021 14:36:39	U	SMPL
HighStd - 1	HighStd	51465	9.6588	-7471B	PPB	16 Nov 2021 14:38:55	U	SMPL
ChkStd - 1	ChkStd	36381	6.8187	-7471B	PPB	16 Nov 2021 14:41:10	U	SMPL
PB140827BL - 1	PBS	-32	-0.0372	-7471B	PPB	16 Nov 2021 14:43:27	U	SMPL
PB140827BS - 1	LCSS	19671	3.6726	-7471B	PPB	16 Nov 2021 14:45:43	U	SMPL
M4695-02 - 1	218	3230	0.5770	-7471B	PPB	16 Nov 2021 14:48:00	U	SMPL
M4695-02DUP - 1	218DUP	2943	0.5230	-7471B	PPB	16 Nov 2021 14:50:16	U	SMPL
M4695-02MS - 1	218MS	22521	4.2092	-7471B	PPB	16 Nov 2021 14:52:33	U	SMPL
M4695-02MSD - 1	218MSD	23810	4.4519	-7471B	PPB	16 Nov 2021 14:57:28	U	SMPL
M4695-04 - 1	289	4339	0.7858	-7471B	PPB	16 Nov 2021 14:59:43	U	SMPL
CCV76 - 1	CCV76	26225	4.9066	-7471B	PPB	16 Nov 2021 15:01:58	U	SMPL
CCB76 - 1	CCB76	5	-0.0302	-7471B	PPB	16 Nov 2021 15:04:13	U	SMPL
M4695-06 - 1	351	3851	0.6939	-7471B	PPB	16 Nov 2021 15:06:28	U	SMPL
M4700-01 - 1	IMPOUND-LOT-SP-JC-11-15-21	120259	22.6114	-7471B	PPB	16 Nov 2021 15:08:44	U	SMPL
M4701-01 - 1	OR-03-111521	3438	0.6162	-7471B	PPB	16 Nov 2021 15:10:59	U	SMPL
M4695-02LX5 - 1		619	0.0854	-7471B	PPB	16 Nov 2021 15:13:17	U	SMPL
M4695-02A - 1		22662	4.2357	-7471B	PPB	16 Nov 2021 15:15:33	U	SMPL
M4700-01DLX5 - 1		25567	4.7827	-7471B	PPB	16 Nov 2021 15:17:49	U	SMPL
CCV77 - 1	CCV77	26224	4.9064	-7471B	PPB	16 Nov 2021 15:20:07	U	SMPL
CCB77 - 1	CCB77	23	-0.0268	-7471B	PPB	16 Nov 2021 15:22:24	U	SMPL