


LB09411

245.1

INSTRUMENT ID: CV1

μ Abs.:

33452



Linear

A= 0.0000e+000

B= 2.9964e-004

C= -4.3268e-002

Rho= 0.9998932

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	✓(ol)
0.0	0.000	-0.005	-0.005	128	0.000	128					14
0.05	0.050	0.057	0.007	333	0.0 %	333	0				8
0.2	0.200	0.217	0.017	869	0.0 %	869					-3
2.5	2.500	2.413	-0.087	8196	0.0 %	8196					2
5.0	5.000	5.109	0.109	17194	0.0 %	17194					0
7.5	7.500	7.480	-0.020	25108	0.0 %	25108					0
10.0	10.000	9.980	-0.020	33452	0.0 %	33452					0

LB117411 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Nov 2021 12:24:00

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	128	-	0.0000	245.1	PPB	17 Nov 2021 12:39:24	S	Std
0.05 - 1	50.05	333	-	0.0500	245.1	PPB	17 Nov 2021 12:41:42	S	Std
0.2 - 1	50.2	869	-	0.2000	245.1	PPB	17 Nov 2021 12:43:59	S	Std
2.5 - 1	52.5	8196	-	2.5000	245.1	PPB	17 Nov 2021 12:46:15	S	Std
5.0 - 1	55	17194	-	5.0000	245.1	PPB	17 Nov 2021 12:48:32	S	Std
7.5 - 1	57.5	25108	-	7.5000	245.1	PPB	17 Nov 2021 12:50:48	S	Std
10.0 - 1	510	33452	-	10.0000	245.1	PPB	17 Nov 2021 12:53:04	S	Std
ICV85 - 1	ICV85	13238	3.9233	-245.1		PPB	17 Nov 2021 12:55:52	U	SMPL
ICB85 - 1	ICB85	121	-0.0070	-245.1		PPB	17 Nov 2021 12:58:07	U	SMPL
CCV75 - 1	CCV75	16985	5.0460	-245.1		PPB	17 Nov 2021 13:00:22	U	SMPL
CCB75 - 1	CCB75	10	-0.0403	-245.1		PPB	17 Nov 2021 13:02:37	U	SMPL
CRA - 1	CRA	325	0.0541	-245.1		PPB	17 Nov 2021 13:04:52	U	SMPL
HighStd - 1	HighStd	34659	10.3418	-245.1		PPB	17 Nov 2021 13:07:07	U	SMPL
ChkStd - 1	ChkStd	24699	7.3574	-245.1		PPB	17 Nov 2021 13:11:58	U	SMPL
PB140826BL - 1	PBW	92	-0.0157	-245.1		PPB	17 Nov 2021 13:14:13	U	SMPL
PB140826BS - 1	LCSW	13225	3.9194	-245.1		PPB	17 Nov 2021 13:16:29	U	SMPL
M4689-03 - 1	TI-604	255	0.0331	-245.1		PPB	17 Nov 2021 13:18:45	U	SMPL
M4689-03DUP - 1	TI-604DUP	280	0.0406	-245.1		PPB	17 Nov 2021 13:21:02	U	SMPL
M4689-03MS - 1	TI-604MS	13535	4.0123	-245.1		PPB	17 Nov 2021 13:23:18	U	SMPL
M4689-03MSD - 1	TI-604MSD	13192	3.9095	-245.1		PPB	17 Nov 2021 13:25:35	U	SMPL
M4690-03 - 1	OB-722	46	-0.0295	-245.1		PPB	17 Nov 2021 13:27:50	U	SMPL
CCV76 - 1	CCV76	17627	5.2384	-245.1		PPB	17 Nov 2021 13:30:06	U	SMPL
CCB76 - 1	CCB76	-5	-0.0448	-245.1		PPB	17 Nov 2021 13:32:21	U	SMPL
M4691-03 - 1	HP-627	238	0.0280	-245.1		PPB	17 Nov 2021 13:34:36	U	SMPL
M4692-01 - 1	WW-EFF	148	0.0011	-245.1		PPB	17 Nov 2021 13:36:52	U	SMPL
M4697-07 - 1	14B-(1-4)-COMP	1567	0.4263	-245.1		PPB	17 Nov 2021 13:39:07	U	SMPL
M4703-07 - 1	14B-(1-4)-COMP	2594	0.7340	-245.1		PPB	17 Nov 2021 13:41:23	U	SMPL
M4689-03LX5 - 1		135	-0.0028	-245.1		PPB	17 Nov 2021 13:43:38	U	SMPL
M4689-03A - 1		13086	3.8778	-245.1		PPB	17 Nov 2021 13:45:54	U	SMPL
CCV77 - 1	CCV77	17289	5.1371	-245.1		PPB	17 Nov 2021 13:48:11	U	SMPL
CCB77 - 1	CCB77	117	-0.0082	-245.1		PPB	17 Nov 2021 13:50:27	U	SMPL