



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 11/18/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 11/17/2021 16:00 DATE & TIME OUT1: 11/18/2021 08:00OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 11/18/2021 08:30 DATE & TIME OUT2: 11/18/2021 10:00Run Number: LB117414BalanceID: WC SC-6OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Dish + Sample Weight (g)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result %
1	LB117414BL	LB117414BL	74.8397	74.8397	74.8397	74.8397	74.8397	74.8397	0.0000	0
2	M4731-01	C-1600-FLOOD-WALL-JG	55.8883	55.8883	75.9421	70.1171	70.1171	70.1171	14.2288	71
3	M4731-01DUP	C-1600-FLOOD-WALL-JGDUP	58.3188	58.3188	80.1371	73.8593	73.8593	73.8593	15.5405	71.2

A = Final Empty Dish Weight (g)

B = Dish + Sample Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result \%} = (C - A) * 100 / (B - A)$$

WORKLIST(Hardcopy Internal Chain)

AB 117614

WorkList Name : TS M4731 S WorkList ID : 154434 Department : Wet-Chemistry Date : 11-17-2021 13:10:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/22/2021	Solid	M4731-01	TS	Cool 4 deg C	TULL02	D32	C-1600-FLOOD-WALL-JC	11/17/2021	SM2540 B

Date/Time 11-17-21 15:00
 Raw Sample Received by: TP (WC)
 Raw Sample Relinquished by: JDCSM

Date/Time 11-17-21 16:30
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: TP (WC)